

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cabarrus Instrument Location Cabarrus County SD
Instrument Serial No. 008590 30 Corban Ave, Concord
704-920-3000

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hult
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590
Test Date: 10/18/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911
Subject's Sex: Male

Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HUTCHINSON, JOSEPH E
Permit Number: 19951E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG016501
Exp Date: 06/14/2012


Test	g/210L	Time
DIAG	Pass	12:37pm
AIR BLK	.00	12:38pm
ACCY CHK	.08	12:39pm
AIR BLK	.00	12:39pm
SUB TEST	.00	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	.00	12:43pm
AIR BLK	.00	12:43pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590 Test Record Number: 1258
Test Date: 10/18/2010 Test Time: 12:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:46pm
FLO	Pass	12:46pm
FC	Pass	12:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:47pm
SRC	Pass	12:47pm
DET	Pass	12:47pm
BAR	Pass	12:47pm
BT	Pass	12:47pm

Blank Tests

Test	Status	Time
AIR	Pass	12:47pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:47pm

CRC Tests

Test	Status	Time
COMP	Pass	12:47pm
CAL	Pass	12:47pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gaston Instrument Location Gaston County SD

Instrument Serial No. 008643 425 N. Marietta Street, Gastonia
704-869-6800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hutchinson
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: 008643
Test Date: 10/11/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920401

Exp Date: 07/23/2011


Test	g/210L	Time
DIAG	Pass	3:58pm
AIR BLK	.00	3:59pm
ACCY CHK	.08	3:59pm
AIR BLK	.00	4:00pm
SUB TEST	.00	4:01pm
AIR BLK	.00	4:02pm
SUB TEST	.00	4:05pm
AIR BLK	.00	4:05pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: 008643 Test Record Number: 1052
Test Date: 10/11/2010 Test Time: 4:07pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:07pm
FLO	Pass	4:07pm
FC	Pass	4:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:07pm
SRC	Pass	4:07pm
DET	Pass	4:07pm
BAR	Pass	4:07pm
BT	Pass	4:07pm

Blank Tests

Test	Status	Time
AIR	Pass	4:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:08pm

CRC Tests

Test	Status	Time
COMP	Pass	4:08pm
CAL	Pass	4:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Cabarrus Instrument Location Cabarrus County SD

Instrument Serial No. 008625 30 Corban Ave SE, Concord
704-920-3000

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hutchinson
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008625
Test Date: 10/05/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG003402


Exp Date: 02/03/2012

Test	g/210L	Time
DIAG	Pass	1:54pm
AIR BLK	.00	1:55pm
ACCY CHK	.08	1:56pm
AIR BLK	.00	1:57pm
SUB TEST	.00	1:57pm
AIR BLK	.00	1:58pm
SUB TEST	.00	2:00pm
AIR BLK	.00	2:00pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008625 Test Record Number: 2153
Test Date: 10/05/2010 Test Time: 2:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:02pm
FLO	Pass	2:02pm
FC	Pass	2:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:02pm
SRC	Pass	2:02pm
DET	Pass	2:02pm
BAR	Pass	2:02pm
BT	Pass	2:02pm

Blank Tests

Test	Status	Time
AIR	Pass	2:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:03pm

CRC Tests

Test	Status	Time
COMP	Pass	2:03pm
CAL	Pass	2:03pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Madison Instrument Location Marcs Hill RD

Instrument Serial No. 008582 Marcs Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MADISON COUNTY MARS HILL PD 560

Serial Number: 008582
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J
Permit Number: 11304E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG920401
Exp Date: 07/23/2011

Test	g/210L	Time
DIAG	Pass	2:11pm
AIR BLK	.00	2:12pm
ACCY CHK	.07	2:13pm
AIR BLK	.00	2:14pm
SUB TEST	.00	2:14pm
AIR BLK	.00	2:15pm
SUB TEST	.00	2:17pm
AIR BLK	.00	2:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MADISON COUNTY MARS HILL PD 560

Serial Number: 008582 Test Record Number: 731
Test Date: 10/06/2010 Test Time: 2:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:18pm
FLO	Pass	2:18pm
FC	Pass	2:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:18pm
SRC	Pass	2:18pm
DET	Pass	2:18pm
BAR	Pass	2:18pm
BT	Pass	2:18pm

Blank Tests

Test	Status	Time
AIR	Pass	2:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:19pm

CRC Tests

Test	Status	Time
COMP	Pass	2:19pm
CAL	Pass	2:19pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Madison Instrument Location Mars Hill P.D.

Instrument Serial No. 008599 Mars Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MADISON COUNTY MARS HILL PD 560

Serial Number: 008599
Test Date: 10/06/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920401

Exp Date: 07/23/2011

Test	g/210L	Time
DIAG	Pass	2:12pm
AIR BLK	.00	2:13pm
ACCY CHK	.08	2:13pm
AIR BLK	.00	2:14pm
SUB TEST	.00	2:14pm
AIR BLK	.00	2:15pm
SUB TEST	.00	2:17pm
AIR BLK	.00	2:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MADISON COUNTY MARS HILL PD 560

Serial Number: 008599 Test Record Number: 342
Test Date: 10/06/2010 Test Time: 2:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:19pm
FLO	Pass	2:19pm
FC	Pass	2:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:19pm
SRC	Pass	2:19pm
DET	Pass	2:19pm
BAR	Pass	2:19pm
BT	Pass	2:19pm

Blank Tests

Test	Status	Time
AIR	Pass	2:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:19pm

CRC Tests

Test	Status	Time
COMP	Pass	2:20pm
CAL	Pass	2:20pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Henderson Instrument Location Henderson Co. Detention

Instrument Serial No. 00 Hendersonville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DETENTION 440

Serial Number: 008822
Test Date: 10/11/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J
Permit Number: 11304E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002602
Exp Date: 01/26/2012

Test	g/210L	Time
DIAG	Pass	6:41pm
AIR BLK	.00	6:42pm
ACCY CHK	.08	6:43pm
AIR BLK	.00	6:44pm
SUB TEST	.00	6:44pm
AIR BLK	.00	6:45pm
SUB TEST	.00	6:46pm
AIR BLK	.00	6:47pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DETENTION 440

Serial Number: 008822 Test Record Number: 965
Test Date: 10/11/2010 Test Time: 6:49pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:49pm
FLO	Pass	6:49pm
FC	Pass	6:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:49pm
SRC	Pass	6:49pm
DET	Pass	6:49pm
BAR	Pass	6:49pm
BT	Pass	6:49pm

Blank Tests

Test	Status	Time
AIR	Pass	6:50pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:50pm

CRC Tests

Test	Status	Time
COMP	Pass	6:50pm
CAL	Pass	6:50pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Henderson Instrument Location Henderson Co Detention

Instrument Serial No. 008806 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806
Test Date: 10/11/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG916701


Exp Date: 06/16/2011

Test	g/210L	Time
DIAG	Pass	6:46pm
AIR BLK	.00	6:47pm
ACCY CHK	.08	6:48pm
AIR BLK	.00	6:48pm
SUB TEST	.00	6:49pm
AIR BLK	.00	6:50pm
SUB TEST	.00	6:51pm
AIR BLK	.00	6:52pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806 Test Record Number: 701
Test Date: 10/11/2010 Test Time: 6:57pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:57pm
FLO	Pass	6:57pm
FC	Pass	6:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:58pm
SRC	Pass	6:58pm
DET	Pass	6:58pm
BAR	Pass	6:58pm
BT	Pass	6:58pm

Blank Tests

Test	Status	Time
AIR	Pass	6:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:58pm

CRC Tests

Test	Status	Time
COMP	Pass	6:58pm
CAL	Pass	6:58pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mitchell Instrument Location Spruce Pine PD

Instrument Serial No. 008726 Spruce Pine, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726
Test Date: 10/20/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002704

Exp Date: 01/27/2012

Test	g/210L	Time
DIAG	Pass	5:09pm
AIR BLK	.00	5:10pm
ACCY CHK	.08	5:10pm
AIR BLK	.00	5:11pm
SUB TEST	.00	5:11pm
AIR BLK	.00	5:12pm
SUB TEST	.00	5:14pm
AIR BLK	.00	5:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726 Test Record Number: 341
Test Date: 10/20/2010 Test Time: 5:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:18pm
FLO	Pass	5:18pm
FC	Pass	5:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:19pm
SRC	Pass	5:19pm
DET	Pass	5:19pm
BAR	Pass	5:19pm
BT	Pass	5:19pm

Blank Tests

Test	Status	Time
AIR	Pass	5:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:19pm

CRC Tests

Test	Status	Time
COMP	Pass	5:19pm
CAL	Pass	5:19pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Caldwell Instrument Location Caldwell Co. Jail

Instrument Serial No. 008803 Lenoir, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CALDWELL COUNTY CALDWELL COUNTY JAIL
130

Serial Number: 008803
Test Date: 10/21/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920401


Exp Date: 07/23/2011

Test	g/210L	Time
DIAG	Pass	3:21pm
AIR BLK	.00	3:22pm
ACCY CHK	.08	3:23pm
AIR BLK	.00	3:24pm
SUB TEST	.00	3:24pm
AIR BLK	.00	3:25pm
SUB TEST	.00	3:27pm
AIR BLK	.00	3:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CALDWELL COUNTY CALDWELL COUNTY JAIL 130

Serial Number: 008803 Test Record Number: 223
Test Date: 10/21/2010 Test Time: 3:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:29pm
FLO	Pass	3:29pm
FC	Pass	3:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:29pm
SRC	Pass	3:29pm
DET	Pass	3:29pm
BAR	Pass	3:29pm
BT	Pass	3:29pm

Blank Tests

Test	Status	Time
AIR	Pass	3:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:30pm

CRC Tests

Test	Status	Time
COMP	Pass	3:30pm
CAL	Pass	3:30pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Yancey Instrument Location Yancey Co. Jail

Instrument Serial No. 008653 Burnsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653
Test Date: 10/25/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002704

Exp Date: 01/27/2012

Test	g/210L	Time
DIAG	Pass	3:44pm
AIR BLK	.00	3:45pm
ACCY CHK	.08	3:46pm
AIR BLK	.00	3:46pm
SUB TEST	.00	3:47pm
AIR BLK	.00	3:48pm
SUB TEST	.00	3:49pm
AIR BLK	.00	3:50pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653 Test Record Number: 706
Test Date: 10/25/2010 Test Time: 3:51pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:51pm
FLO	Pass	3:51pm
FC	Pass	3:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:52pm
SRC	Pass	3:52pm
DET	Pass	3:52pm
BAR	Pass	3:52pm
BT	Pass	3:52pm

Blank Tests

Test	Status	Time
AIR	Pass	3:52pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:52pm

CRC Tests

Test	Status	Time
COMP	Pass	3:52pm
CAL	Pass	3:52pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Craven Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008898 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008898
Test Date: 10/01/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:23pm
AIR BLK	.00	11:24pm
ACCY CHK	.07	11:25pm
AIR BLK	.00	11:25pm
SUB TEST	.00	11:26pm
AIR BLK	.00	11:27pm
SUB TEST	.00	11:28pm
AIR BLK	.00	11:29pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008898 Test Record Number: 502
Test Date: 10/02/2010 Test Time: 12:28am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:29am
FLO	Pass	12:29am
FC	Pass	12:29am

Temperature Tests

Test	Status	Time
FC1	Pass	12:29am
SRC	Pass	12:29am
DET	Pass	12:29am
BAR	Pass	12:29am
BT	Pass	12:29am

Blank Tests

Test	Status	Time
AIR	Pass	12:30am

Printer Tests

Test	Status	Time
PRNT	Pass	12:30am

CRC Tests

Test	Status	Time
COMP	Pass	12:30am
CAL	Pass	12:30am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County CITRUVEN Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008939 SAP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

6001
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008939
Test Date: 10/01/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:24pm
AIR BLK	.00	11:25pm
ACCY CHK	.08	11:26pm
AIR BLK	.00	11:27pm
SUB TEST	.00	11:27pm
AIR BLK	.00	11:28pm
SUB TEST	.00	11:30pm
AIR BLK	.00	11:31pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008939 Test Record Number: 443
Test Date: 10/01/2010 Test Time: 11:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:36pm
FLO	Pass	11:36pm
FC	Pass	11:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:36pm
SRC	Pass	11:36pm
DET	Pass	11:36pm
BAR	Pass	11:36pm
BT	Pass	11:36pm

Blank Tests

Test	Status	Time
AIR	Pass	11:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:37pm

CRC Tests

Test	Status	Time
COMP	Pass	11:37pm
CAL	Pass	11:37pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Craven Instrument Location BAT mobile unit 6

Instrument Serial No. 008869 SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. C. Moore
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008869 Test Record Number: 367
Test Date: 10/02/2010 Test Time: 2:14am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:14am
FLO	Pass	2:14am
FC	Pass	2:14am

Temperature Tests

Test	Status	Time
FC1	Pass	2:14am
SRC	Pass	2:14am
DET	Pass	2:14am
BAR	Pass	2:14am
BT	Pass	2:14am

Blank Tests

Test	Status	Time
AIR	Pass	2:15am

Printer Tests

Test	Status	Time
PRNT	Pass	2:15am

CRC Tests

Test	Status	Time
COMP	Pass	2:15am
CAL	Pass	2:15am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 6 240

Serial Number: 008869
Test Date: 10/02/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	2:06am
AIR BLK	.00	2:07am
ACCY CHK	.08	2:07am
AIR BLK	.00	2:09am
SUB TEST	.00	2:09am
AIR BLK	.00	2:10am
SUB TEST	.00	2:12am
AIR BLK	.00	2:13am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Lenoir Instrument Location BAT mobile unit 6

Instrument Serial No. 608939 Kinston

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY BAT MOBILE UNIT 6 530

Serial Number: 008939 Test Record Number: 447
Test Date: 10/02/2010 Test Time: 11:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:12pm
FLO	Pass	11:12pm
FC	Pass	11:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:12pm
SRC	Pass	11:12pm
DET	Pass	11:12pm
BAR	Pass	11:12pm
BT	Pass	11:12pm

Blank Tests

Test	Status	Time
AIR	Pass	11:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:13pm

CRC Tests

Test	Status	Time
COMP	Pass	11:13pm
CAL	Pass	11:13pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

LENOIR COUNTY BAT MOBILE UNIT 6 530

Serial Number: 008939

Test Date: 10/02/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:04pm
AIR BLK	.00	11:05pm
ACCY CHK	.08	11:06pm
AIR BLK	.00	11:06pm
SUB TEST	.00	11:07pm
AIR BLK	.00	11:08pm
SUB TEST	.00	11:09pm
AIR BLK	.00	11:10pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Carteret Instrument Location BAT Mobile Unit #6

Instrument Serial No. 008939 Atlantic Beach

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008939 Test Record Number: 451
Test Date: 10/10/2010 Test Time: 12:02am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:02am
FLO	Pass	12:02am
FC	Pass	12:02am

Temperature Tests

Test	Status	Time
FC1	Pass	12:02am
SRC	Pass	12:02am
DET	Pass	12:02am
BAR	Pass	12:02am
BT	Pass	12:02am

Blank Tests

Test	Status	Time
AIR	Pass	12:03am

Printer Tests

Test	Status	Time
PRNT	Pass	12:03am

CRC Tests

Test	Status	Time
COMP	Pass	12:03am
CAL	Pass	12:03am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008939
Test Date: 10/09/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:54pm
AIR BLK	.00	11:55pm
ACCY CHK	.08	11:56pm
AIR BLK	.00	11:57pm
SUB TEST	.00	11:57pm
AIR BLK	.00	11:58pm
SUB TEST	.00	11:59pm
AIR BLK	.00	12:00am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location BAT mobile unit
Instrument Serial No. 008869 S.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Moore
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008869 Test Record Number: 370
Test Date: 10/30/2010 Test Time: 11:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:29pm
FLO	Pass	11:29pm
FC	Pass	11:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:29pm
SRC	Pass	11:29pm
DET	Pass	11:29pm
BAR	Pass	11:29pm
BT	Pass	11:29pm

Blank Tests

Test	Status	Time
AIR	Pass	11:30pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:30pm

CRC Tests

Test	Status	Time
COMP	Pass	11:30pm
CAL	Pass	11:30pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008869
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name: PREVENTIVE, MAINTANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:20pm
AIR BLK	.00	11:21pm
ACCY CHK	.08	11:21pm
AIR BLK	.00	11:22pm
SUB TEST	.00	11:23pm
AIR BLK	.00	11:24pm
SUB TEST	.00	11:25pm
AIR BLK	.00	11:26pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location BAT Mobile Unit 46
Instrument Serial No. 008898 S.O

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008898 Test Record Number: 512
Test Date: 10/30/2010 Test Time: 11:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30pm
FLO	Pass	11:30pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:30pm
SRC	Pass	11:30pm
DET	Pass	11:30pm
BAR	Pass	11:30pm
BT	Pass	11:30pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:31pm

CRC Tests

Test	Status	Time
COMP	Pass	11:31pm
CAL	Pass	11:31pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008898
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:21pm
AIR BLK	.00	11:22pm
ACCY CHK	.07	11:22pm
AIR BLK	.00	11:23pm
SUB TEST	.00	11:24pm
AIR BLK	.00	11:24pm
SUB TEST	.00	11:26pm
AIR BLK	.00	11:27pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location BAT Mobile unit #6
Instrument Serial No. 008939 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of OCT, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008939 Test Record Number: 461
Test Date: 10/30/2010 Test Time: 11:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30pm
FLO	Pass	11:30pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:30pm
SRC	Pass	11:30pm
DET	Pass	11:30pm
BAR	Pass	11:30pm
BT	Pass	11:30pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:31pm

CRC Tests

Test	Status	Time
COMP	Pass	11:31pm
CAL	Pass	11:31pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 6 090

Serial Number: 008939
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:20pm
AIR BLK	.00	11:21pm
ACCY CHK	.08	11:21pm
AIR BLK	.00	11:22pm
SUB TEST	.00	11:24pm
AIR BLK	.00	11:25pm
SUB TEST	.00	11:26pm
AIR BLK	.00	11:27pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Onslow Instrument Location BAT Mobile unit 6
Instrument Serial No. 008898 JACKSONVILLE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCT, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Proctor
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY BAT MOBILE UNIT 6 660

Serial Number: 008898 Test Record Number: 508
Test Date: 10/29/2010 Test Time: 7:00pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:00pm
FLO	Pass	7:00pm
FC	Pass	7:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:01pm
SRC	Pass	7:01pm
DET	Pass	7:01pm
BAR	Pass	7:01pm
BT	Pass	7:01pm

Blank Tests

Test	Status	Time
AIR	Pass	7:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:01pm

CRC Tests

Test	Status	Time
COMP	Pass	7:01pm
CAL	Pass	7:01pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 6 660

Serial Number: 008898
Test Date: 10/29/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	6:52pm
AIR BLK	.00	6:53pm
ACCY CHK	.07	6:54pm
AIR BLK	.00	6:54pm
SUB TEST	.00	6:55pm
AIR BLK	.00	6:56pm
SUB TEST	.00	6:57pm
AIR BLK	.00	6:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Danlow Instrument Location BAT Mobile unit 6
Instrument Serial No. 008939 JACKSONVILLE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCT, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Proctor
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY BAT MOBILE UNIT 6 660

Serial Number: 008939 Test Record Number: 456
Test Date: 10/29/2010 Test Time: 6:58pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:58pm
FLO	Pass	6:58pm
FC	Pass	6:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:59pm
SRC	Pass	6:59pm
DET	Pass	6:59pm
BAR	Pass	6:59pm
BT	Pass	6:59pm

Blank Tests

Test	Status	Time
AIR	Pass	6:59pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:59pm

CRC Tests

Test	Status	Time
COMP	Pass	6:59pm
CAL	Pass	6:59pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

ONslow COUNTY BAT MOBILE UNIT 6 660

Serial Number: 008939

Test Date: 10/29/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	6:51pm
AIR BLK	.00	6:52pm
ACCY CHK	.08	6:52pm
AIR BLK	.00	6:53pm
SUB TEST	.00	6:54pm
AIR BLK	.00	6:55pm
SUB TEST	.00	6:56pm
AIR BLK	.00	6:57pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Carteret Instrument Location BAT mobile unit 6

Instrument Serial No. 008898 Morehead City

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of November, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008898 Test Record Number: 522
Test Date: 11/05/2010 Test Time: 9:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:14pm
FLO	Pass	9:14pm
FC	Pass	9:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:14pm
SRC	Pass	9:14pm
DET	Pass	9:14pm
BAR	Pass	9:14pm
BT	Pass	9:14pm

Blank Tests

Test	Status	Time
AIR	Pass	9:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:14pm

CRC Tests

Test	Status	Time
COMP	Pass	9:15pm
CAL	Pass	9:15pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008898
Test Date: 11/05/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	9:06pm
AIR BLK	.00	9:07pm
ACCY CHK	.07	9:07pm
AIR BLK	.00	9:08pm
SUB TEST	.00	9:08pm
AIR BLK	.00	9:09pm
SUB TEST	.00	9:11pm
AIR BLK	.00	9:12pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret Instrument Location BAT Mobile unit 76
Instrument Serial No. 008939 Morehead City

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of November, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008939 Test Record Number: 472
Test Date: 11/05/2010 Test Time: 9:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:18pm
FLO	Pass	9:18pm
FC	Pass	9:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:19pm
SRC	Pass	9:19pm
DET	Pass	9:19pm
BAR	Pass	9:19pm
BT	Pass	9:19pm

Blank Tests

Test	Status	Time
AIR	Pass	9:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:19pm

CRC Tests

Test	Status	Time
COMP	Pass	9:19pm
CAL	Pass	9:19pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

CARTERET COUNTY BAT MOBILE UNIT 6 150

Serial Number: 008939
Test Date: 11/05/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	9:06pm
AIR BLK	.00	9:07pm
ACCY CHK	.08	9:08pm
AIR BLK	.00	9:09pm
SUB TEST	.00	9:09pm
AIR BLK	.00	9:10pm
SUB TEST	.00	9:12pm
AIR BLK	.00	9:13pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location BAT Mobile unit 4
Instrument Serial No. 008939 Greenville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008939 Test Record Number: 470
Test Date: 10/31/2010 Test Time: 5:42pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:42pm
FLO	Pass	5:42pm
FC	Pass	5:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:42pm
SRC	Pass	5:42pm
DET	Pass	5:42pm
BAR	Pass	5:42pm
BT	Pass	5:42pm

Blank Tests

Test	Status	Time
AIR	Pass	5:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:43pm

CRC Tests

Test	Status	Time
COMP	Pass	5:43pm
CAL	Pass	5:43pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008939
Test Date: 10/31/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	5:33pm
AIR BLK	.00	5:34pm
ACCY CHK	.08	5:34pm
AIR BLK	.00	5:35pm
SUB TEST	.00	5:36pm
AIR BLK	.00	5:36pm
SUB TEST	.00	5:39pm
AIR BLK	.00	5:40pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location BA7 Mobile unit 4
Instrument Serial No. 008898 Greenville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Moore
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008898 Test Record Number: 517
Test Date: 10/31/2010 Test Time: 5:42pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:42pm
FLO	Pass	5:42pm
FC	Pass	5:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:42pm
SRC	Pass	5:42pm
DET	Pass	5:42pm
BAR	Pass	5:42pm
BT	Pass	5:42pm

Blank Tests

Test	Status	Time
AIR	Pass	5:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:43pm

CRC Tests

Test	Status	Time
COMP	Pass	5:43pm
CAL	Pass	5:43pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008898
Test Date: 10/31/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	5:34pm
AIR BLK	.00	5:35pm
ACCY CHK	.07	5:36pm
AIR BLK	.00	5:37pm
SUB TEST	.00	5:37pm
AIR BLK	.00	5:38pm
SUB TEST	.00	5:40pm
AIR BLK	.00	5:41pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Pitt Instrument Location BAT Mobile Unit 6

Instrument Serial No. 008869 Greenville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



A. C. Maden
Signature of Certifying Official

601 []
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008869 Test Record Number: 373
Test Date: 10/31/2010 Test Time: 5:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:40pm
FLO	Pass	5:40pm
FC	Pass	5:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:41pm
SRC	Pass	5:41pm
DET	Pass	5:41pm
BAR	Pass	5:41pm
BT	Pass	5:41pm

Blank Tests

Test	Status	Time
AIR	Pass	5:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:41pm

CRC Tests

Test	Status	Time
COMP	Pass	5:41pm
CAL	Pass	5:41pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008869
Test Date: 10/31/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	5:31pm
AIR BLK	.00	5:32pm
ACCY CHK	.07	5:33pm
AIR BLK	.00	5:34pm
SUB TEST	.00	5:34pm
AIR BLK	.00	5:35pm
SUB TEST	.00	5:37pm
AIR BLK	.00	5:38pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location Pasquotank Co. Public Safety Bldg
Instrument Serial No. 008588 200 E. Colonial Ave., Elizabeth City, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda Reed
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY PUBLIC SAFETY BLDG
690

Serial Number: 008588
Test Date: 10/27/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA
Permit Number: 11646E
Effective:
09/01/2009-09/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925103
Exp Date: 09/08/2011


Test	g/210L	Time
DIAG	Pass	12:37pm
AIR BLK	.00	12:38pm
ACCY CHK	.08	12:38pm
AIR BLK	.00	12:39pm
SUB TEST	.00	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	.00	12:42pm
AIR BLK	.00	12:43pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008588 Test Record Number: 536
Test Date: 10/27/2010 Test Time: 12:45pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:45pm
FLO	Pass	12:45pm
FC	Pass	12:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:45pm
SRC	Pass	12:45pm
DET	Pass	12:45pm
BAR	Pass	12:45pm
BT	Pass	12:45pm

Blank Tests

Test	Status	Time
AIR	Pass	12:46pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:46pm

CRC Tests

Test	Status	Time
COMP	Pass	12:46pm
CAL	Pass	12:46pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Bertie Instrument Location Windsor PD, 2nd Precinct
Instrument Serial No. 004897 122 E. Granville, Windsor, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Lincoln Reed
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BERTIE COUNTY WINDSOR PD 2ND PRCNT 070

Serial Number: 008897
Test Date: 10/07/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA A
Permit Number: 11646E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG920303
Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	12:11pm
AIR BLK	.00	12:12pm
ACCY CHK	.08	12:12pm
AIR BLK	.00	12:13pm
SUB TEST	.00	12:14pm
AIR BLK	.00	12:14pm
SUB TEST	.00	12:16pm
AIR BLK	.00	12:17pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BERTIE COUNTY WINDSOR PD 2ND PRCNT 070

Serial Number: 008897 Test Record Number: 387
Test Date: 10/07/2010 Test Time: 12:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:20pm
SRC	Pass	12:20pm
DET	Pass	12:20pm
BAR	Pass	12:20pm
BT	Pass	12:20pm

Blank Tests

Test	Status	Time
AIR	Pass	12:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:20pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Dare Instrument Location Kill Devil Hills P.O.

Instrument Serial No. 009844 107 Town Hall Dr., Kill Devil Hills, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5th day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda Keese
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844

Test Date: 10/05/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG925103

Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	12:19pm
AIR BLK	.00	12:20pm
ACCY CHK	.08	12:21pm
AIR BLK	.00	12:22pm
SUB TEST	.00	12:22pm
AIR BLK	.00	12:23pm
SUB TEST	.00	12:25pm
AIR BLK	.00	12:26pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844 Test Record Number: 690
Test Date: 10/05/2010 Test Time: 12:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:27pm
FLO	Pass	12:27pm
FC	Pass	12:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:27pm
SRC	Pass	12:27pm
DET	Pass	12:27pm
BAR	Pass	12:27pm
BT	Pass	12:27pm

Blank Tests

Test	Status	Time
AIR	Pass	12:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:28pm

CRC Tests

Test	Status	Time
COMP	Pass	12:28pm
CAL	Pass	12:28pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. 008646 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of December, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	9:47am
AIR BLK	.00	9:48am
ACCY CHK	.07	9:49am
AIR BLK	.00	9:50am
SUB TEST	.00	9:50am
AIR BLK	.00	9:51am
SUB TEST	.00	9:53am
AIR BLK	.00	9:54am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646 Test Record Number: 992
Test Date: 10/06/2010 Test Time: 9:55am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55am
FLO	Pass	9:55am
FC	Pass	9:55am

Temperature Tests

Test	Status	Time
FC1	Pass	9:55am
SRC	Pass	9:55am
DET	Pass	9:55am
BAR	Pass	9:55am
BT	Pass	9:55am

Blank Tests

Test	Status	Time
AIR	Pass	9:56am


Printer Tests

Test	Status	Time
PRNT	Pass	9:56am

CRC Tests

Test	Status	Time
COMP	Pass	9:56am
CAL	Pass	9:56am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. DD9668 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of December, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012


Test	g/210L	Time
DIAG	Pass	9:36am
AIR BLK	.00	9:37am
ACCY CHK	.07	9:37am
AIR BLK	.00	9:38am
SUB TEST	.00	9:39am
AIR BLK	.00	9:40am
SUB TEST	.00	9:41am
AIR BLK	.00	9:42am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668 Test Record Number: 1411
Test Date: 10/06/2010 Test Time: 9:43am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:44am
FLO	Pass	9:44am
FC	Pass	9:44am

Temperature Tests

Test	Status	Time
FC1	Pass	9:44am
SRC	Pass	9:44am
DET	Pass	9:44am
BAR	Pass	9:44am
BT	Pass	9:44am

Blank Tests

Test	Status	Time
AIR	Pass	9:44am


Printer Tests

Test	Status	Time
PRNT	Pass	9:44am

CRC Tests

Test	Status	Time
COMP	Pass	9:45am
CAL	Pass	9:45am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. 008662 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Lois A. Hesse
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008662
Test Date: 10/06/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG916602

Exp Date: 06/15/2011


Test	g/210L	Time
DIAG	Pass	9:50am
AIR BLK	.00	9:50am
ACCY CHK	.07	9:51am
AIR BLK	.00	9:52am
SUB TEST	.00	9:53am
AIR BLK	.00	9:54am
SUB TEST	.00	9:55am
AIR BLK	.00	9:56am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008662 Test Record Number: 469
Test Date: 10/06/2010 Test Time: 10:05am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:05am
FLO	Pass	10:05am
FC	Pass	10:05am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:06am


Printer Tests

Test	Status	Time
PRNT	Pass	10:06am

CRC Tests

Test	Status	Time
COMP	Pass	10:07am
CAL	Pass	10:07am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County LEE Instrument Location SANFORD POLICE

Instrument Serial No. 008867 DEPT, SANFORD, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul T. Cannon
Signature of Certifying Official

578
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867
Test Date: 10/28/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SIMMONS, PAUL T
Permit Number: 08619E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG825401
Exp Date: 10/02/2011

Test	g/210L	Time
DIAG	Pass	4:52pm
AIR BLK	.00	4:53pm
ACCY CHK	.07	4:54pm
AIR BLK	.00	4:55pm
SUB TEST	.00	4:55pm
AIR BLK	.00	4:56pm
SUB TEST	.00	4:57pm
AIR BLK	.00	4:58pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867 Test Record Number: 313
Test Date: 10/28/2010 Test Time: 4:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:59pm
FLO	Pass	4:59pm
FC	Pass	4:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:00pm
SRC	Pass	5:00pm
DET	Pass	5:00pm
BAR	Pass	5:00pm
BT	Pass	5:00pm

Blank Tests

Test	Status	Time
AIR	Pass	5:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:00pm

CRC Tests

Test	Status	Time
COMP	Pass	5:00pm
CAL	Pass	5:00pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Hyde Instrument Location Hyde Co. Sheriff's Office
Instrument Serial No. 008801 1233 Main St., Swan Quarter, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801
Test Date: 10/26/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG916602

Exp Date: 06/15/2011

Test	g/210L	Time
DIAG	Pass	11:08am
AIR BLK	.00	11:09am
ACCY CHK	.07	11:09am
AIR BLK	.00	11:10am
SUB TEST	.00	11:11am
AIR BLK	.00	11:11am
SUB TEST	.00	11:13am
AIR BLK	.00	11:14am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801 Test Record Number: 173
Test Date: 10/26/2010 Test Time: 11:15am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:16am
FLO	Pass	11:16am
FC	Pass	11:16am

Temperature Tests

Test	Status	Time
FC1	Pass	11:16am
SRC	Pass	11:16am
DET	Pass	11:16am
BAR	Pass	11:16am
BT	Pass	11:16am

Blank Tests

Test	Status	Time
AIR	Pass	11:16am

Printer Tests

Test	Status	Time
PRNT	Pass	11:16am

CRC Tests

Test	Status	Time
COMP	Pass	11:17am
CAL	Pass	11:17am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Martin Instrument Location Martin Co. S.O.

Instrument Serial No. 008879 305 E. Main St., Williamston, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. [Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008879
Test Date: 10/25/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012


Test	g/210L	Time
DIAG	Pass	9:58am
AIR BLK	.00	9:59am
ACCY CHK	.08	10:00am
AIR BLK	.00	10:01am
SUB TEST	.00	10:01am
AIR BLK	.00	10:02am
SUB TEST	.00	10:04am
AIR BLK	.00	10:05am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008879 Test Record Number: 223
Test Date: 10/25/2010 Test Time: 10:06am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:06am
FLO	Pass	10:06am
FC	Pass	10:06am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:07am


Printer Tests

Test	Status	Time
PRNT	Pass	10:07am

CRC Tests

Test	Status	Time
COMP	Pass	10:07am
CAL	Pass	10:07am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Lenoir Instrument Location Kinston P.D.

Instrument Serial No. 008624 205 E. King St., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. P.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624
Test Date: 10/22/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG916602

Exp Date: 06/15/2011

Test	g/210L	Time
DIAG	Pass	11:28am
AIR BLK	.00	11:29am
ACCY CHK	.08	11:29am
AIR BLK	.00	11:30am
SUB TEST	.00	11:31am
AIR BLK	.00	11:31am
SUB TEST	.00	11:33am
AIR BLK	.00	11:34am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624 Test Record Number: 780
Test Date: 10/22/2010 Test Time: 11:36am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:36am
FLO	Pass	11:36am
FC	Pass	11:36am

Temperature Tests

Test	Status	Time
FC1	Pass	11:36am
SRC	Pass	11:36am
DET	Pass	11:36am
BAR	Pass	11:36am
BT	Pass	11:36am

Blank Tests

Test	Status	Time
AIR	Pass	11:37am

Printer Tests

Test	Status	Time
PRNT	Pass	11:37am

CRC Tests

Test	Status	Time
COMP	Pass	11:37am
CAL	Pass	11:37am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chowan Instrument Location Chowan Co Public Safety Cen
Instrument Serial No. 008895 305 W. Freemason St, Edenton, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895
Test Date: 10/21/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920303

Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	11:09am
AIR BLK	.00	11:10am
ACCY CHK	.08	11:11am
AIR BLK	.00	11:12am
SUB TEST	.00	11:12am
AIR BLK	.00	11:13am
SUB TEST	.00	11:15am
AIR BLK	.00	11:16am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895 Test Record Number: 307
Test Date: 10/21/2010 Test Time: 11:17am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:17am
FLO	Pass	11:17am
FC	Pass	11:18am

Temperature Tests

Test	Status	Time
FC1	Pass	11:18am
SRC	Pass	11:18am
DET	Pass	11:18am
BAR	Pass	11:18am
BT	Pass	11:18am

Blank Tests

Test	Status	Time
AIR	Pass	11:18am

Printer Tests

Test	Status	Time
PRNT	Pass	11:18am

CRC Tests

Test	Status	Time
COMP	Pass	11:19am
CAL	Pass	11:19am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008727 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cuth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008727
Test Date: 10/21/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R
Permit Number: 08457E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG910501
Exp Date: 04/15/2011

Test	g/210L	Time
DIAG	Pass	12:03pm
AIR BLK	.00	12:04pm
ACCY CHK	.08	12:04pm
AIR BLK	.00	12:05pm
SUB TEST	.00	12:06pm
AIR BLK	.00	12:07pm
SUB TEST	.00	12:08pm
AIR BLK	.00	12:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008727 Test Record Number: 476
Test Date: 10/21/2010 Test Time: 12:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:11pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:11pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008723 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723
Test Date: 10/21/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG910501

Exp Date: 04/15/2011

Test	g/210L	Time
DIAG	Pass	12:04pm
AIR BLK	.00	12:05pm
ACCY CHK	.08	12:05pm
AIR BLK	.00	12:06pm
SUB TEST	.00	12:07pm
AIR BLK	.00	12:08pm
SUB TEST	.00	12:09pm
AIR BLK	.00	12:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723 Test Record Number: 241
Test Date: 10/21/2010 Test Time: 12:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Cherokee Instrument Location Cherokee Co. Jail

Instrument Serial No. 008911 Murphy, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Carol R. Guth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHEROKEE COUNTY CHEROKEE COUNTY JAIL
190

Serial Number: 008911
Test Date: 10/25/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R
Permit Number: 08457E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG910501
Exp Date: 04/15/2011

Test	g/210L	Time
DIAG	Pass	12:35pm
AIR BLK	.00	12:36pm
ACCY CHK	.07	12:36pm
AIR BLK	.00	12:37pm
SUB TEST	.00	12:38pm
AIR BLK	.00	12:39pm
SUB TEST	.00	12:40pm
AIR BLK	.00	12:41pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CHEROKEE COUNTY CHEROKEE COUNTY JAIL 190

Serial Number: 008911 Test Record Number: 57
Test Date: 10/25/2010 Test Time: 12:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:32pm
SRC	Pass	12:32pm
DET	Pass	12:32pm
BAR	Pass	12:32pm
BT	Pass	12:32pm

Blank Tests

Test	Status	Time
AIR	Pass	12:33pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:33pm

CRC Tests

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Macon Instrument Location Highlands P.D.

Instrument Serial No. 008795 Highlands, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Guth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MACON COUNTY HIGHLANDS PD 550

Serial Number: 008795
Test Date: 10/28/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R
Permit Number: 08457E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG910501
Exp Date: 04/15/2011

Test	g/210L	Time
DIAG	Pass	12:30pm
AIR BLK	.00	12:31pm
ACCY CHK	.07	12:31pm
AIR BLK	.00	12:32pm
SUB TEST	.00	12:33pm
AIR BLK	.00	12:33pm
SUB TEST	.00	12:35pm
AIR BLK	.00	12:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MACON COUNTY HIGHLANDS PD 550

Serial Number: 008795 Test Record Number: 170
Test Date: 10/28/2010 Test Time: 12:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:37pm
FLO	Pass	12:37pm
FC	Pass	12:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:37pm
SRC	Pass	12:37pm
DET	Pass	12:37pm
BAR	Pass	12:37pm
BT	Pass	12:37pm

Blank Tests

Test	Status	Time
AIR	Pass	12:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:38pm

CRC Tests

Test	Status	Time
COMP	Pass	12:38pm
CAL	Pass	12:38pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 THOMASVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: 008707
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	10:27pm
AIR BLK	.00	10:28pm
ACCY CHK	.08	10:29pm
AIR BLK	.00	10:30pm
SUB TEST	.00	10:30pm
AIR BLK	.00	10:31pm
SUB TEST	.00	10:33pm
AIR BLK	.00	10:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: 008707 Test Record Number: 706
Test Date: 10/30/2010 Test Time: 10:34pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:34pm
FLO	Pass	10:34pm
FC	Pass	10:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:34pm
SRC	Pass	10:34pm
DET	Pass	10:34pm
BAR	Pass	10:34pm
BT	Pass	10:34pm

Blank Tests

Test	Status	Time
AIR	Pass	10:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:35pm

CRC Tests

Test	Status	Time
COMP	Pass	10:35pm
CAL	Pass	10:35pm

Preventive Maintenance
Status: Pass

Oliver Ray Barnes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DAVIDSON Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 THOMASVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: 008647
Test Date: 10/30/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011703

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	10:42pm
AIR BLK	.00	10:43pm
ACCY CHK	.08	10:44pm
AIR BLK	.00	10:45pm
SUB TEST	.00	10:45pm
AIR BLK	.00	10:46pm
SUB TEST	.00	10:48pm
AIR BLK	.00	10:49pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY BAT MOBILE UNIT 3 280

Serial Number: 008647 Test Record Number: 949
Test Date: 10/30/2010 Test Time: 10:51pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51pm
FLO	Pass	10:51pm
FC	Pass	10:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:51pm
SRC	Pass	10:51pm
DET	Pass	10:51pm
BAR	Pass	10:51pm
BT	Pass	10:51pm

Blank Tests

Test	Status	Time
AIR	Pass	10:52pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:52pm

CRC Tests

Test	Status	Time
COMP	Pass	10:52pm
CAL	Pass	10:52pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County GUILFORD Instrument Location BAT MOBILE UNIT 3
Instrument Serial No. 008707 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008707
Test Date: 10/29/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:19pm
AIR BLK	.00	11:20pm
ACCY CHK	.08	11:21pm
AIR BLK	.00	11:21pm
SUB TEST	.00	11:22pm
AIR BLK	.00	11:23pm
SUB TEST	.00	11:25pm
AIR BLK	.00	11:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008707 Test Record Number: 701
Test Date: 10/29/2010 Test Time: 11:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:26pm
FLO	Pass	11:26pm
FC	Pass	11:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:27pm
SRC	Pass	11:27pm
DET	Pass	11:27pm
BAR	Pass	11:27pm
BT	Pass	11:27pm

Blank Tests

Test	Status	Time
AIR	Pass	11:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:27pm

CRC Tests

Test	Status	Time
COMP	Pass	11:27pm
CAL	Pass	11:27pm

Preventive Maintenance
Status: Pass

Alexis Kay Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County GUILFORD Instrument Location BAT MOBILE UNIT 3
Instrument Serial No. 008647 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Aileen Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008647
Test Date: 10/29/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011703
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:24pm
AIR BLK	.00	11:25pm
ACCY CHK	.08	11:25pm
AIR BLK	.00	11:26pm
SUB TEST	.00	11:27pm
AIR BLK	.00	11:27pm
SUB TEST	.00	11:29pm
AIR BLK	.00	11:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008647 Test Record Number: 942
Test Date: 10/29/2010 Test Time: 11:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:31pm
FLO	Pass	11:31pm
FC	Pass	11:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:31pm
SRC	Pass	11:31pm
DET	Pass	11:31pm
BAR	Pass	11:31pm
BT	Pass	11:31pm

Blank Tests

Test	Status	Time
AIR	Pass	11:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:32pm

CRC Tests

Test	Status	Time
COMP	Pass	11:32pm
CAL	Pass	11:32pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County GUILFORD Instrument Location BAT MOBILE UNIT 3
Instrument Serial No. 008736 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Allen Ray Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008736
Test Date: 10/29/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG925102

Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	11:33pm
AIR BLK	.00	11:34pm
ACCY CHK	.07	11:35pm
AIR BLK	.00	11:36pm
SUB TEST	.00	11:36pm
AIR BLK	.00	11:37pm
SUB TEST	.00	11:39pm
AIR BLK	.00	11:40pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008736 Test Record Number: 222
Test Date: 10/29/2010 Test Time: 11:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40pm
FLO	Pass	11:40pm
FC	Pass	11:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:41pm
SRC	Pass	11:41pm
DET	Pass	11:41pm
BAR	Pass	11:41pm
BT	Pass	11:41pm

Blank Tests

Test	Status	Time
AIR	Pass	11:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:41pm

CRC Tests

Test	Status	Time
COMP	Pass	11:41pm
CAL	Pass	11:41pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County PITT Instrument Location BAT Mobile Unit #5

Instrument Serial No. 008600 Grenville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31st day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008600 Test Record Number: 705
Test Date: 10/31/2010 Test Time: 10:25pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:26pm
FLO	Pass	10:26pm
FC	Pass	10:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:26pm
SRC	Pass	10:26pm
DET	Pass	10:26pm
BAR	Pass	10:26pm
BT	Pass	10:26pm

Blank Tests

Test	Status	Time
AIR	Pass	10:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:27pm

CRC Tests

Test	Status	Time
COMP	Pass	10:27pm
CAL	Pass	10:27pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008600
Test Date: 10/31/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920302

Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	10:16pm
AIR BLK	.00	10:17pm
ACCY CHK	.08	10:18pm
AIR BLK	.00	10:19pm
SUB TEST	.00	10:20pm
AIR BLK	.00	10:20pm
SUB TEST	.00	10:23pm
AIR BLK	.00	10:23pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County PITT Instrument Location BAT MOBILE UNIT #5
Instrument Serial No. 008698 GARLANDVILLE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31ST day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008698 Test Record Number: 585
Test Date: 10/31/2010 Test Time: 10:27pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:27pm
FLO	Pass	10:27pm
FC	Pass	10:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:27pm
SRC	Pass	10:27pm
DET	Pass	10:27pm
BAR	Pass	10:27pm
BT	Pass	10:27pm

Blank Tests

Test	Status	Time
AIR	Pass	10:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:28pm

CRC Tests

Test	Status	Time
COMP	Pass	10:28pm
CAL	Pass	10:28pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008698
Test Date: 10/31/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	10:17pm
AIR BLK	.00	10:18pm
ACCY CHK	.08	10:19pm
AIR BLK	.00	10:20pm
SUB TEST	.00	10:20pm
AIR BLK	.00	10:21pm
SUB TEST	.00	10:23pm
AIR BLK	.00	10:24pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location BAT MOBILE UNIT #5
Instrument Serial No. 008788 GREENVILLE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31st day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official 636 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008788 Test Record Number: 484
Test Date: 10/31/2010 Test Time: 10:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:28pm
FLO	Pass	10:28pm
FC	Pass	10:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:29pm
SRC	Pass	10:29pm
DET	Pass	10:29pm
BAR	Pass	10:29pm
BT	Pass	10:29pm

Blank Tests

Test	Status	Time
AIR	Pass	10:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:29pm

CRC Tests

Test	Status	Time
COMP	Pass	10:29pm
CAL	Pass	10:29pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

PITT COUNTY BAT MOBILE UNIT 5 730

Serial Number: 008788

Test Date: 10/31/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904903


Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	10:19pm
AIR BLK	.00	10:20pm
ACCY CHK	.07	10:21pm
AIR BLK	.00	10:22pm
SUB TEST	.00	10:22pm
AIR BLK	.00	10:23pm
SUB TEST	.00	10:25pm
AIR BLK	.00	10:26pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gaillard Instrument Location BAT mobile unit #5

Instrument Serial No. 028600

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Steph G. Moberly 656
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008600 Test Record Number: 700
Test Date: 10/30/2010 Test Time: 10:05pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:05pm
FLO	Pass	10:05pm
FC	Pass	10:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:05pm
SRC	Pass	10:05pm
DET	Pass	10:05pm
BAR	Pass	10:05pm
BT	Pass	10:05pm

Blank Tests

Test	Status	Time
AIR	Pass	10:06pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:06pm

CRC Tests

Test	Status	Time
COMP	Pass	10:06pm
CAL	Pass	10:06pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008600
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:55pm
AIR BLK	.00	9:56pm
ACCY CHK	.08	9:57pm
AIR BLK	.00	9:57pm
SUB TEST	.00	9:58pm
AIR BLK	.00	9:59pm
SUB TEST	.00	10:01pm
AIR BLK	.00	10:02pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GREENE Instrument Location BAT MOBILE UNIT #3-

Instrument Serial No. 008698

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30TH day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008698 Test Record Number: 581
Test Date: 10/30/2010 Test Time: 10:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:10pm
FLO	Pass	10:10pm
FC	Pass	10:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:10pm
SRC	Pass	10:10pm
DET	Pass	10:10pm
BAR	Pass	10:10pm
BT	Pass	10:10pm

Blank Tests

Test	Status	Time
AIR	Pass	10:11pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:11pm

CRC Tests

Test	Status	Time
COMP	Pass	10:11pm
CAL	Pass	10:11pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008698
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920302

Exp Date: 07/22/2011


Test	g/210L	Time
DIAG	Pass	10:00pm
AIR BLK	.00	10:01pm
ACCY CHK	.08	10:01pm
AIR BLK	.00	10:02pm
SUB TEST	.00	10:03pm
AIR BLK	.00	10:04pm
SUB TEST	.00	10:05pm
AIR BLK	.00	10:06pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GREENE Instrument Location BAT MOBILE WIT #5

Instrument Serial No. 008758

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] 636
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008788 Test Record Number: 482
Test Date: 10/30/2010 Test Time: 10:21pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:21pm
FLO	Pass	10:21pm
FC	Pass	10:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:21pm
SRC	Pass	10:21pm
DET	Pass	10:21pm
BAR	Pass	10:21pm
BT	Pass	10:21pm

Blank Tests

Test	Status	Time
AIR	Pass	10:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:22pm

CRC Tests

Test	Status	Time
COMP	Pass	10:22pm
CAL	Pass	10:22pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

GREENE COUNTY BAT MOBILE UNIT 5 390

Serial Number: 008788
Test Date: 10/30/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904903

Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	10:03pm
AIR BLK	.00	10:04pm
ACCY CHK	.07	10:04pm
AIR BLK	.00	10:05pm
SUB TEST	.00	10:08pm
AIR BLK	.00	10:09pm
SUB TEST	.00	10:11pm
AIR BLK	.00	10:11pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 8600 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28TH day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Steph O. Thomas
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008600 Test Record Number: 684
Test Date: 10/28/2010 Test Time: 10:04pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:04pm
FLO	Pass	10:04pm
FC	Pass	10:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:04pm
SRC	Pass	10:04pm
DET	Pass	10:04pm
BAR	Pass	10:04pm
BT	Pass	10:04pm

Blank Tests

Test	Status	Time
AIR	Pass	10:05pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:05pm

CRC Tests

Test	Status	Time
COMP	Pass	10:05pm
CAL	Pass	10:05pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008600
Test Date: 10/28/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:52pm
AIR BLK	.00	9:53pm
ACCY CHK	.08	9:53pm
AIR BLK	.00	9:54pm
SUB TEST	.00	9:55pm
AIR BLK	.00	9:56pm
SUB TEST	.00	9:58pm
AIR BLK	.00	9:59pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location BAT MOBILE UNIT #5
Instrument Serial No. 08698 Rolco

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698 Test Record Number: 567
Test Date: 10/28/2010 Test Time: 10:21pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:22pm
FLO	Pass	10:22pm
FC	Pass	10:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:22pm
SRC	Pass	10:22pm
DET	Pass	10:22pm
BAR	Pass	10:22pm
BT	Pass	10:22pm

Blank Tests

Test	Status	Time
AIR	Pass	10:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:23pm

CRC Tests

Test	Status	Time
COMP	Pass	10:23pm
CAL	Pass	10:23pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698

Test Date: 10/28/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920302

Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:54pm
AIR BLK	.00	9:55pm
ACCY CHK	.08	9:55pm
AIR BLK	.00	9:56pm
SUB TEST	.00	9:57pm
AIR BLK	.00	9:58pm
SUB TEST	.00	10:00pm
AIR BLK	.00	10:00pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 Raleigh

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy; 6
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28TH day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stephen G. Thompson 636
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788 Test Record Number: 470
Test Date: 10/28/2010 Test Time: 10:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13pm
FLO	Pass	10:14pm
FC	Pass	10:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:14pm
SRC	Pass	10:14pm
DET	Pass	10:14pm
BAR	Pass	10:14pm
BT	Pass	10:14pm

Blank Tests

Test	Status	Time
AIR	Pass	10:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:14pm

CRC Tests

Test	Status	Time
COMP	Pass	10:15pm
CAL	Pass	10:15pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test . . .

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788
Test Date: 10/28/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904903


Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	9:56pm
AIR BLK	.00	9:57pm
ACCY CHK	.08	9:58pm
AIR BLK	.00	9:59pm
SUB TEST	.00	9:59pm
AIR BLK	.00	10:00pm
SUB TEST	.00	10:02pm
AIR BLK	.00	10:03pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location BATMOBILE Unit #5

Instrument Serial No. 008698 Wendell

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stuart G. III

Signature of Certifying Official

636

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698 Test Record Number: 578
Test Date: 10/29/2010 Test Time: 10:20pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:20pm
FLO	Pass	10:20pm
FC	Pass	10:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:20pm
SRC	Pass	10:20pm
DET	Pass	10:20pm
BAR	Pass	10:20pm
BT	Pass	10:20pm

Blank Tests

Test	Status	Time
AIR	Pass	10:21pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:21pm

CRC Tests

Test	Status	Time
COMP	Pass	10:21pm
CAL	Pass	10:21pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698
Test Date: 10/29/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	10:11pm
AIR BLK	.00	10:12pm
ACCY CHK	.08	10:13pm
AIR BLK	.00	10:13pm
SUB TEST	.00	10:14pm
AIR BLK	.00	10:15pm
SUB TEST	.00	10:17pm
AIR BLK	.00	10:17pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location BAT MOBILE UNIT #5

Instrument Serial No. 008788 WENDALL

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29TH day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stephen B. Thibault
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788 Test Record Number: 478
Test Date: 10/29/2010 Test Time: 10:21pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:21pm
FLO	Pass	10:21pm
FC	Pass	10:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:22pm
SRC	Pass	10:22pm
DET	Pass	10:22pm
BAR	Pass	10:22pm
BT	Pass	10:22pm

Blank Tests

Test	Status	Time
AIR	Pass	10:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:22pm

CRC Tests

Test	Status	Time
COMP	Pass	10:22pm
CAL	Pass	10:22pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788

Test Date: 10/29/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG904903

Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	10:13pm
AIR BLK	.00	10:14pm
ACCY CHK	.07	10:14pm
AIR BLK	.00	10:15pm
SUB TEST	.00	10:16pm
AIR BLK	.00	10:17pm
SUB TEST	.00	10:18pm
AIR BLK	.00	10:19pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location WAKE FOREST P.D.

Instrument Serial No. 008700 401 OWEN ST. WAKE FOREST, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700
Test Date: 10/28/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG916701
Exp Date: 06/16/2011


Test	g/210L	Time
DIAG	Pass	12:40pm
AIR BLK	.00	12:41pm
ACCY CHK	.07	12:41pm
AIR BLK	.00	12:42pm
SUB TEST	.00	12:43pm
AIR BLK	.00	12:44pm
SUB TEST	.00	12:45pm
AIR BLK	.00	12:46pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700 Test Record Number: 411
Test Date: 10/28/2010 Test Time: 12:47pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:47pm
FLO	Pass	12:47pm
FC	Pass	12:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:48pm
SRC	Pass	12:48pm
DET	Pass	12:48pm
BAR	Pass	12:48pm
BT	Pass	12:48pm

Blank Tests

Test	Status	Time
AIR	Pass	12:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:48pm

CRC Tests

Test	Status	Time
COMP	Pass	12:49pm
CAL	Pass	12:49pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CCBI

Instrument Serial No. 008615 330 S. SAUSURRY ST., RALEIGH, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CCBI 910

Serial Number: 008615
Test Date: 10/29/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011701

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:33am
AIR BLK	.00	11:34am
ACCY CHK	.08	11:35am
AIR BLK	.00	11:36am
SUB TEST	.00	11:36am
AIR BLK	.00	11:37am
SUB TEST	.00	11:39am
AIR BLK	.00	11:40am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CCBI 910

Serial Number: 008615 Test Record Number: 2121
Test Date: 10/29/2010 Test Time: 11:41am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:41am
FLO	Pass	11:41am
FC	Pass	11:41am

Temperature Tests

Test	Status	Time
FC1	Pass	11:42am
SRC	Pass	11:42am
DET	Pass	11:42am
BAR	Pass	11:42am
BT	Pass	11:42am

Blank Tests

Test	Status	Time
AIR	Pass	11:42am

Printer Tests

Test	Status	Time
PRNT	Pass	11:42am

CRC Tests

Test	Status	Time
COMP	Pass	11:42am
CAL	Pass	11:42am

Preventive Maintenance
Status: Pass

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CCRC

Instrument Serial No. 008826 330 S. SALISBURY ST. RALEIGH, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CCBI 910

Serial Number: 008826
Test Date: 10/29/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG016501

Exp Date: 06/14/2012

Test	g/210L	Time
DIAG	Pass	11:27am
AIR BLK	.00	11:28am
ACCY CHK	.08	11:29am
AIR BLK	.00	11:30am
SUB TEST	.00	11:31am
AIR BLK	.00	11:32am
SUB TEST	.00	11:33am
AIR BLK	.00	11:34am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CCBI 910

Serial Number: 008826 Test Record Number: 3629
Test Date: 10/29/2010 Test Time: 11:37am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:37am
FLO	Pass	11:37am
FC	Pass	11:37am

Temperature Tests

Test	Status	Time
FC1	Pass	11:37am
SRC	Pass	11:37am
DET	Pass	11:37am
BAR	Pass	11:37am
BT	Pass	11:37am

Blank Tests

Test	Status	Time
AIR	Pass	11:38am

Printer Tests

Test	Status	Time
PRNT	Pass	11:38am

CRC Tests

Test	Status	Time
COMP	Pass	11:38am
CAL	Pass	11:38am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WARREN

Instrument Location WARREN CO. JAIL


Instrument Serial No. 008793

Hwy 58 WARRENTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793
Test Date: 10/27/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG902603
Exp Date: 01/26/2011

Test	g/210L	Time
DIAG	Pass	11:41am
AIR BLK	.00	11:41am
ACCY CHK	.08	11:42am
AIR BLK	.00	11:43am
SUB TEST	.00	11:43am
AIR BLK	.00	11:44am
SUB TEST	.00	11:46am
AIR BLK	.00	11:47am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793 Test Record Number: 386
Test Date: 10/27/2010 Test Time: 11:48am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:48am
FLO	Pass	11:48am
FC	Pass	11:48am

Temperature Tests

Test	Status	Time
FC1	Pass	11:48am
SRC	Pass	11:48am
DET	Pass	11:48am
BAR	Pass	11:48am
BT	Pass	11:48am

Blank Tests

Test	Status	Time
AIR	Pass	11:49am

Printer Tests

Test	Status	Time
PRNT	Pass	11:49am

CRC Tests

Test	Status	Time
COMP	Pass	11:49am
CAL	Pass	11:49am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WARREN Instrument Location NORLINA P.D.

Instrument Serial No. 008945 101 MAIN ST. NORLINA, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945
Test Date: 10/27/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG910601
Exp Date: 04/16/2011

Test	g/210L	Time
DIAG	Pass	11:14am
AIR BLK	.00	11:15am
ACCY CHK	.07	11:16am
AIR BLK	.00	11:17am
SUB TEST	.00	11:17am
AIR BLK	.00	11:18am
SUB TEST	.00	11:20am
AIR BLK	.00	11:20am

Reported AC: .00 g/210L



Signature of Chemical Analyst
Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945 Test Record Number: 184
Test Date: 10/27/2010 Test Time: 11:22am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:23am
FLO	Pass	11:23am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:23am

Printer Tests

Test	Status	Time
PRNT	Pass	11:23am

CRC Tests

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County VANCE Instrument Location VANCE CO. SHERIFFS DEPT
Instrument Serial No. 008937 156 CHURCH ST. HENDERSON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937
Test Date: 10/27/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG901901


Exp Date: 01/19/2011

Test	g/210L	Time
DIAG	Pass	10:29am
AIR BLK	.00	10:30am
ACCY CHK	.08	10:30am
AIR BLK	.00	10:31am
SUB TEST	.00	10:32am
AIR BLK	.00	10:33am
SUB TEST	.00	10:34am
AIR BLK	.00	10:35am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937 Test Record Number: 839
Test Date: 10/27/2010 Test Time: 10:36am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:36am
FLO	Pass	10:36am
FC	Pass	10:36am

Temperature Tests

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

Blank Tests

Test	Status	Time
AIR	Pass	10:37am

Printer Tests

Test	Status	Time
PRNT	Pass	10:37am

CRC Tests

Test	Status	Time
COMP	Pass	10:37am
CAL	Pass	10:37am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County VANCE Instrument Location VANCE CO SHERIFFS DEPT

Instrument Serial No. 008870 156 CHURCH ST. HENDERSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870
Test Date: 10/27/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902603

Exp Date: 01/26/2011


Test	g/210L	Time
DIAG	Pass	10:32am
AIR BLK	.00	10:33am
ACCY CHK	.08	10:33am
AIR BLK	.00	10:34am
SUB TEST	.00	10:35am
AIR BLK	.00	10:36am
SUB TEST	.00	10:37am
AIR BLK	.00	10:38am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870 Test Record Number: 240
Test Date: 10/27/2010 Test Time: 10:39am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:39am
FLO	Pass	10:39am
FC	Pass	10:39am

Temperature Tests

Test	Status	Time
FC1	Pass	10:40am
SRC	Pass	10:40am
DET	Pass	10:40am
BAR	Pass	10:40am
BT	Pass	10:40am

Blank Tests

Test	Status	Time
AIR	Pass	10:40am

Printer Tests

Test	Status	Time
PRNT	Pass	10:40am

CRC Tests

Test	Status	Time
COMP	Pass	10:40am
CAL	Pass	10:40am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford


Instrument Location High Point Jail

Instrument Serial No. 008655

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008655

Test Date: 10/14/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	10:37am
AIR BLK	.00	10:38am
ACCY CHK	.08	10:39am
AIR BLK	.00	10:39am
SUB TEST	.00	10:40am
AIR BLK	.00	10:41am
SUB TEST	.00	10:42am
AIR BLK	.00	10:43am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008655 Test Record Number: 1407
Test Date: 10/14/2010 Test Time: 10:44am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:44am
FLO	Pass	10:44am
FC	Pass	10:44am

Temperature Tests

Test	Status	Time
FC1	Pass	10:44am
SRC	Pass	10:44am
DET	Pass	10:44am
BAR	Pass	10:44am
BT	Pass	10:44am

Blank Tests

Test	Status	Time
AIR	Pass	10:45am


Printer Tests

Test	Status	Time
PRNT	Pass	10:45am

CRC Tests

Test	Status	Time
COMP	Pass	10:45am
CAL	Pass	10:45am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location High Point Jail

Instrument Serial No. 008718

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008718

Test Date: 10/14/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012


Test	g/210L	Time
DIAG	Pass	10:42am
AIR BLK	.00	10:42am
ACCY CHK	.07	10:43am
AIR BLK	.00	10:44am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:47am
AIR BLK	.00	10:48am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008718 Test Record Number: 481
Test Date: 10/14/2010 Test Time: 10:48am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:49am
FLO	Pass	10:49am
FC	Pass	10:49am

Temperature Tests

Test	Status	Time
FC1	Pass	10:49am
SRC	Pass	10:49am
DET	Pass	10:49am
BAR	Pass	10:49am
BT	Pass	10:49am

Blank Tests

Test	Status	Time
AIR	Pass	10:50am


Printer Tests

Test	Status	Time
PRNT	Pass	10:50am

CRC Tests

Test	Status	Time
COMP	Pass	10:50am
CAL	Pass	10:50am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Archdale Police
Instrument Serial No. 008791 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791
Test Date: 10/20/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG904902
Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	1:11pm
AIR BLK	.00	1:12pm
ACCY CHK	.08	1:12pm
AIR BLK	.00	1:13pm
SUB TEST	.00	1:14pm
AIR BLK	.00	1:15pm
SUB TEST	.00	1:16pm
AIR BLK	.00	1:17pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791 Test Record Number: 433
Test Date: 10/20/2010 Test Time: 1:24pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:25pm
SRC	Pass	1:25pm
DET	Pass	1:25pm
BAR	Pass	1:25pm
BT	Pass	1:25pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location Greensboro Police

Instrument Serial No. 008725 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725
Test Date: 10/14/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG011702
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	2:52pm
AIR BLK	.00	2:53pm
ACCY CHK	.08	2:54pm
AIR BLK	.00	2:55pm
SUB TEST	.00	2:55pm
AIR BLK	.00	2:56pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:59pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725 Test Record Number: 1580
Test Date: 10/14/2010 Test Time: 2:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:00pm
FLO	Pass	3:00pm
FC	Pass	3:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:00pm
SRC	Pass	3:00pm
DET	Pass	3:00pm
BAR	Pass	3:00pm
BT	Pass	3:00pm

Blank Tests

Test	Status	Time
AIR	Pass	3:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:00pm

CRC Tests

Test	Status	Time
COMP	Pass	3:01pm
CAL	Pass	3:01pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cleveland Instrument Location Cleveland County S.D.
Instrument Serial No. 008927 100 Justice Pl., Shelby
704-484-4888

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby O. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND COUNTY SD
220

Serial Number: 008927
Test Date: 10/27/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925103
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	10:02am
AIR BLK	.00	10:03am
ACCY CHK	.08	10:03am
AIR BLK	.00	10:04am
SUB TEST	.00	10:05am
AIR BLK	.00	10:06am
SUB TEST	.00	10:07am
AIR BLK	.00	10:08am

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND COUNTY SD 220

Serial Number: 008927 Test Record Number: 356
Test Date: 10/27/2010 Test Time: 10:09am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:09am
FLO	Pass	10:09am
FC	Pass	10:09am

Temperature Tests

Test	Status	Time
FC1	Pass	10:09am
SRC	Pass	10:09am
DET	Pass	10:09am
BAR	Pass	10:09am
BT	Pass	10:09am

Blank Tests

Test	Status	Time
AIR	Pass	10:10am

Printer Tests

Test	Status	Time
PRNT	Pass	10:10am

CRC Tests

Test	Status	Time
COMP	Pass	10:10am
CAL	Pass	10:10am

Preventive Maintenance
Status: Pass

Boley D. Willis
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Catawba Instrument Location Hickory P.D.

Instrument Serial No. 008841 347 2nd Ave. SW, Hickory
828-324-2060

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Rolley E. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841
Test Date: 10/13/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925103
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	12:50pm
AIR BLK	.00	12:51pm
ACCY CHK	.07	12:51pm
AIR BLK	.00	12:52pm
SUB TEST	.00	12:55pm
AIR BLK	.00	12:56pm
SUB TEST	.00	12:57pm
AIR BLK	.00	12:58pm

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841 Test Record Number: 750
Test Date: 10/13/2010 Test Time: 1:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:02pm
FLO	Pass	1:02pm
FC	Pass	1:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:02pm
SRC	Pass	1:02pm
DET	Pass	1:02pm
BAR	Pass	1:02pm
BT	Pass	1:02pm

Blank Tests

Test	Status	Time
AIR	Pass	1:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:03pm

CRC Tests

Test	Status	Time
COMP	Pass	1:03pm
CAL	Pass	1:03pm

Preventive Maintenance
Status: Pass

Bobby D. Willis
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Catawba County S. D.
Instrument Serial No. 008821 100 B Southwest Blvd., Newton
828-464-5241

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby E. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821
Test Date: 10/13/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG902603
Exp Date: 01/26/2011

Test	g/210L	Time
DIAG	Pass	10:50am
AIR BLK	.00	10:51am
ACCY CHK	.07	10:52am
AIR BLK	.00	10:52am
SUB TEST	.00	10:53am
AIR BLK	.00	10:54am
SUB TEST	.00	10:55am
AIR BLK	.00	10:56am

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821 Test Record Number: 584
Test Date: 10/13/2010 Test Time: 10:57am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:57am
FLO	Pass	10:57am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:58am

Printer Tests

Test	Status	Time
PRNT	Pass	10:58am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance
Status: Pass

Boleyn D. Willis
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Catawba Instrument Location Catawba County S. D.

Instrument Serial No. 008687 100 B Southwest Blvd, Newton
828-464-5241

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby E. Willie
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008687
Test Date: 10/13/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG003401
Exp Date: 02/03/2012

Test	g/210L	Time
DIAG	Pass	10:20am
AIR BLK	.00	10:20am
ACCY CHK	.08	10:21am
AIR BLK	.00	10:22am
SUB TEST	.00	10:22am
AIR BLK	.00	10:23am
SUB TEST	.00	10:25am
AIR BLK	.00	10:26am

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008687 Test Record Number: 1037
Test Date: 10/13/2010 Test Time: 10:27am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27am
FLO	Pass	10:27am
FC	Pass	10:27am

Temperature Tests

Test	Status	Time
FC1	Pass	10:27am
SRC	Pass	10:27am
DET	Pass	10:27am
BAR	Pass	10:27am
BT	Pass	10:27am

Blank Tests

Test	Status	Time
AIR	Pass	10:28am

Printer Tests

Test	Status	Time
PRNT	Pass	10:28am

CRC Tests

Test	Status	Time
COMP	Pass	10:28am
CAL	Pass	10:28am

Preventive Maintenance
Status: Pass

Bolley D. Willis
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Iredell Instrument Location Mooreville P.D.

Instrument Serial No. 008694 750 W. Iredell Ave., Mooreville
704-664-3311

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby L. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: 008694
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG003402
Exp Date: 02/03/2012

Test	g/210L	Time
DIAG	Pass	12:43pm
AIR BLK	.00	12:43pm
ACCY CHK	.08	12:44pm
AIR BLK	.00	12:45pm
SUB TEST	.00	12:46pm
AIR BLK	.00	12:46pm
SUB TEST	.00	12:48pm
AIR BLK	.00	12:49pm

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: 008694 Test Record Number: 121
Test Date: 10/06/2010 Test Time: 12:50pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:50pm
FLO	Pass	12:50pm
FC	Pass	12:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:50pm
SRC	Pass	12:50pm
DET	Pass	12:50pm
BAR	Pass	12:50pm
BT	Pass	12:50pm

Blank Tests

Test	Status	Time
AIR	Pass	12:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:51pm

CRC Tests

Test	Status	Time
COMP	Pass	12:51pm
CAL	Pass	12:51pm

Preventive Maintenance
Status: Pass

Bobby D. Willis
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rutherford Instrument Location Forest City P. D.
Instrument Serial No. 008889 187 S. Church St. Forest City
828-245-5555

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby D. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889
Test Date: 10/05/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D
Permit Number: 08010E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925103
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	11:27am
AIR BLK	.00	11:27am
ACCY CHK	.08	11:28am
AIR BLK	.00	11:29am
SUB TEST	.00	11:29am
AIR BLK	.00	11:30am
SUB TEST	.00	11:32am
AIR BLK	.00	11:33am

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst
Court CVR

Bobby D. Willis
Analyst

Intox EC/IR-II: Preventive Maintenance

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889 Test Record Number: 323
Test Date: 10/05/2010 Test Time: 11:34am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:34am

Temperature Tests

Test	Status	Time
FC1	Pass	11:34am
SRC	Pass	11:34am
DET	Pass	11:34am
BAR	Pass	11:34am
BT	Pass	11:34am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am


Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rutherford Instrument Location Rutherford County S.D.
Instrument Serial No. 008914 400 N. Washington St., Rutherfordton
828-627-6247

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bobby D. Willis
Signature of Certifying Official

557
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RUTHERFORD COUNTY RUTHERFORD COUNTY SD
800

Serial Number: 008914
Test Date: 10/05/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: WILLIS, BOBBY D

Permit Number: 08010E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920401

Exp Date: 07/23/2011

Test	g/210L	Time
DIAG	Pass	10:10am
AIR BLK	.00	10:11am
ACCY CHK	.08	10:11am
AIR BLK	.00	10:12am
SUB TEST	.00	10:15am
AIR BLK	.00	10:16am
SUB TEST	.00	10:17am
AIR BLK	.00	10:18am

Reported AC: .00 g/210L

Bobby D. Willis
Signature of Chemical Analyst

Court CVR

Bobby D. Willis
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RUTHERFORD COUNTY RUTHERFORD COUNTY SD 800

Serial Number: 008914 Test Record Number: 580
Test Date: 10/05/2010 Test Time: 10:20am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:21am
FLO	Pass	10:21am
FC	Pass	10:21am

Temperature Tests

Test	Status	Time
FC1	Pass	10:21am
SRC	Pass	10:21am
DET	Pass	10:21am
BAR	Pass	10:21am
BT	Pass	10:21am

Blank Tests

Test	Status	Time
AIR	Pass	10:22am


Printer Tests

Test	Status	Time
PRNT	Pass	10:22am

CRC Tests

Test	Status	Time
COMP	Pass	10:22am
CAL	Pass	10:22am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Yadkin Instrument Location Yadkin Co Jail

Instrument Serial No. 008944

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944
Test Date: 10/25/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: WEAVER, GEORGE A

Permit Number: 09442E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902901

Exp Date: 01/29/2011

Test	g/210L	Time
DIAG	Pass	1:32pm
AIR BLK	.00	1:32pm
ACCY CHK	.08	1:33pm
AIR BLK	.00	1:34pm
SUB TEST	.00	1:34pm
AIR BLK	.00	1:35pm
SUB TEST	.00	1:37pm
AIR BLK	.00	1:38pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944 Test Record Number: 554
Test Date: 10/25/2010 Test Time: 1:38pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:39pm
FLO	Pass	1:39pm
FC	Pass	1:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:39pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:39pm

CRC Tests

Test	Status	Time
COMP	Pass	1:40pm
CAL	Pass	1:40pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Yadkin Instrument Location Yadkinville P.P.

Instrument Serial No. 008925

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKINVILLE PD 980

Serial Number: 008925
Test Date: 10/25/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: WEAVER, GEORGE A

Permit Number: 09442E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902901

Exp Date: 01/29/2011

Test	g/210L	Time
DIAG	Pass	12:43pm
AIR BLK	.00	12:44pm
ACCY CHK	.08	12:44pm
AIR BLK	.00	12:45pm
SUB TEST	.00	12:46pm
AIR BLK	.00	12:47pm
SUB TEST	.00	12:48pm
AIR BLK	.00	12:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKINVILLE PD 980

Serial Number: 008925 Test Record Number: 191
Test Date: 10/25/2010 Test Time: 12:50pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:50pm
FLO	Pass	12:50pm
FC	Pass	12:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:50pm
SRC	Pass	12:50pm
DET	Pass	12:50pm
BAR	Pass	12:50pm
BT	Pass	12:50pm

Blank Tests

Test	Status	Time
AIR	Pass	12:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:51pm

CRC Tests

Test	Status	Time
COMP	Pass	12:51pm
CAL	Pass	12:51pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilkes Instrument Location Wilkes County
Instrument Serial No. 008843 Court House

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO COURTHOUSE 960

Serial Number: 008843
Test Date: 10/22/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: WEAVER, GEORGE A
Permit Number: 09442E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002703
Exp Date: 01/27/2012

Test	g/210L	Time
DIAG	Pass	1:43pm
AIR BLK	.00	1:43pm
ACCY CHK	.07	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	.00	1:45pm
AIR BLK	.00	1:46pm
SUB TEST	.00	1:48pm
AIR BLK	.00	1:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO COURTHOUSE 960

Serial Number: 008843 Test Record Number: 722
Test Date: 10/22/2010 Test Time: 1:49pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:49pm
FLO	Pass	1:49pm
FC	Pass	1:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:50pm
SRC	Pass	1:50pm
DET	Pass	1:50pm
BAR	Pass	1:50pm
BT	Pass	1:50pm

Blank Tests

Test	Status	Time
AIR	Pass	1:50pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:50pm

CRC Tests

Test	Status	Time
COMP	Pass	1:50pm
CAL	Pass	1:50pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Allyhany Instrument Location Allyhany Co Jail
Instrument Serial No. 008890

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: 008890
Test Date: 10/19/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: WEAVER, GEORGE A
Permit Number: 09442E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011702
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	4:48pm
AIR BLK	.00	4:48pm
ACCY CHK	.08	4:49pm
AIR BLK	.00	4:50pm
SUB TEST	.00	4:50pm
AIR BLK	.00	4:51pm
SUB TEST	.00	4:53pm
AIR BLK	.00	4:54pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: 008890 Test Record Number: 227
Test Date: 10/19/2010 Test Time: 4:54pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:55pm
FLO	Pass	4:55pm
FC	Pass	4:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:55pm
SRC	Pass	4:55pm
DET	Pass	4:55pm
BAR	Pass	4:55pm
BT	Pass	4:55pm

Blank Tests

Test	Status	Time
AIR	Pass	4:55pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:55pm

CRC Tests

Test	Status	Time
COMP	Pass	4:56pm
CAL	Pass	4:56pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Ashe Instrument Location Ashe Co Jail
140 GOVERNMENT CIR
Instrument Serial No. 008849 JEFFERSON, NC 28640

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849
Test Date: 10/19/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: WEAVER, GEORGE A
Permit Number: 09442E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011702
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	3:40pm
AIR BLK	.00	3:41pm
ACCY CHK	.08	3:42pm
AIR BLK	.00	3:43pm
SUB TEST	.00	3:44pm
AIR BLK	.00	3:45pm
SUB TEST	.00	3:46pm
AIR BLK	.00	3:47pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849 Test Record Number: 439
Test Date: 10/19/2010 Test Time: 3:48pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:48pm
FLO	Pass	3:48pm
FC	Pass	3:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:48pm
SRC	Pass	3:48pm
DET	Pass	3:48pm
BAR	Pass	3:48pm
BT	Pass	3:48pm

Blank Tests

Test	Status	Time
AIR	Pass	3:49pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:49pm

CRC Tests

Test	Status	Time
COMP	Pass	3:49pm
CAL	Pass	3:49pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Rockingham Co Jail
Instrument Serial No. 008796 Reidsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

632
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL
780

Serial Number: 008796
Test Date: 10/18/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: WEAVER, GEORGE A
Permit Number: 09442E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011703
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:51am
ACCY CHK	.08	11:52am
AIR BLK	.00	11:53am
SUB TEST	.00	11:53am
AIR BLK	.00	11:54am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL 780

Serial Number: 008796 Test Record Number: 656
Test Date: 10/18/2010 Test Time: 11:58am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:59am

Printer Tests

Test	Status	Time
PRNT	Pass	11:59am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County PERSON Instrument Location PERSON Co. LEL

Instrument Serial No. 008880 120 COURT ST. RUXBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880

Test Date: 10/26/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904902


Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	1:02pm
AIR BLK	.00	1:03pm
ACCY CHK	.08	1:04pm
AIR BLK	.00	1:05pm
SUB TEST	.00	1:05pm
AIR BLK	.00	1:06pm
SUB TEST	.00	1:08pm
AIR BLK	.00	1:09pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880 Test Record Number: 304
Test Date: 10/26/2010 Test Time: 1:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:11pm
FLO	Pass	1:11pm
FC	Pass	1:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:11pm
SRC	Pass	1:11pm
DET	Pass	1:11pm
BAR	Pass	1:11pm
BT	Pass	1:11pm

Blank Tests

Test	Status	Time
AIR	Pass	1:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:12pm

CRC Tests

Test	Status	Time
COMP	Pass	1:12pm
CAL	Pass	1:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County PERSON Instrument Location PERSON Co. LLC

Instrument Serial No. 008693 120 COURT ST REXBORO, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693

Test Date: 10/26/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920301

Exp Date: 07/22/2011


Test	g/210L	Time
DIAG	Pass	1:01pm
AIR BLK	.00	1:02pm
ACCY CHK	.08	1:03pm
AIR BLK	.00	1:04pm
SUB TEST	.00	1:04pm
AIR BLK	.00	1:05pm
SUB TEST	.00	1:07pm
AIR BLK	.00	1:08pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693 Test Record Number: 614
Test Date: 10/26/2010 Test Time: 1:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:09pm
FLO	Pass	1:09pm
FC	Pass	1:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:09pm
SRC	Pass	1:09pm
DET	Pass	1:09pm
BAR	Pass	1:09pm
BT	Pass	1:09pm

Blank Tests

Test	Status	Time
AIR	Pass	1:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:10pm

CRC Tests

Test	Status	Time
COMP	Pass	1:10pm
CAL	Pass	1:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County NASH Instrument Location NASH Co. Jail

Instrument Serial No. 008630 NASHVILLE, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: 008630
Test Date: 10/15/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, AINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name:
QUARANTELLLO, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011703
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:05am
AIR BLK	.00	11:06am
ACCY CHK	.08	11:07am
AIR BLK	.00	11:07am
SUB TEST	.00	11:08am
AIR BLK	.00	11:09am
SUB TEST	.00	11:10am
AIR BLK	.00	11:11am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY NASH COUNTY JAIL 630

Serial Number: 008630 Test Record Number: 1818
Test Date: 10/15/2010 Test Time: 11:13am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:13am
FLO	Pass	11:13am
FC	Pass	11:13am

Temperature Tests

Test	Status	Time
FC1	Pass	11:13am
SRC	Pass	11:13am
DET	Pass	11:13am
BAR	Pass	11:13am
BT	Pass	11:13am

Blank Tests

Test	Status	Time
AIR	Pass	11:14am


Printer Tests

Test	Status	Time
PRNT	Pass	11:14am

CRC Tests

Test	Status	Time
COMP	Pass	11:14am
CAL	Pass	11:14am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLINTON P.D.

Instrument Serial No. 008815 #7 W. MASON ST. FRANKLINTON, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: 008815
Test Date: 10/25/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG904903
Exp Date: 02/18/2011


Test	g/210L	Time
DIAG	Pass	11:58am
AIR BLK	.00	11:59am
ACCY CHK	.08	12:00pm
AIR BLK	.00	12:01pm
SUB TEST	.00	12:01pm
AIR BLK	.00	12:02pm
SUB TEST	.00	12:04pm
AIR BLK	.00	12:05pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: 008815 Test Record Number: 429
Test Date: 10/25/2010 Test Time: 12:06pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:07pm
FLO	Pass	12:07pm
FC	Pass	12:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:07pm
SRC	Pass	12:07pm
DET	Pass	12:07pm
BAR	Pass	12:07pm
BT	Pass	12:07pm

Blank Tests

Test	Status	Time
AIR	Pass	12:07pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:07pm

CRC Tests

Test	Status	Time
COMP	Pass	12:08pm
CAL	Pass	12:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLIN CO. JAN

Instrument Serial No. 008933 285 T KEMP RD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933
Test Date: 10/25/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904903

Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	11:00am
ACCY CHK	.08	11:00am
AIR BLK	.00	11:01am
SUB TEST	.00	11:02am
AIR BLK	.00	11:03am
SUB TEST	.00	11:04am
AIR BLK	.00	11:05am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933 Test Record Number: 388
Test Date: 10/25/2010 Test Time: 11:06am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:06am
FLO	Pass	11:06am
FC	Pass	11:07am

Temperature Tests

Test	Status	Time
FC1	Pass	11:07am
SRC	Pass	11:07am
DET	Pass	11:07am
BAR	Pass	11:07am
BT	Pass	11:07am

Blank Tests

Test	Status	Time
AIR	Pass	11:07am


Printer Tests

Test	Status	Time
PRNT	Pass	11:07am

CRC Tests

Test	Status	Time
COMP	Pass	11:08am
CAL	Pass	11:08am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLIN CO. JAIL

Instrument Serial No. 008942 285 T KEMP RD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942
Test Date: 10/25/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG904903
Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	10:53am
AIR BLK	.00	10:54am
ACCY CHK	.08	10:54am
AIR BLK	.00	10:55am
SUB TEST	.00	10:56am
AIR BLK	.00	10:57am
SUB TEST	.00	10:58am
AIR BLK	.00	10:59am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942 Test Record Number: 315
Test Date: 10/25/2010 Test Time: 11:01am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:02am
FLO	Pass	11:02am
FC	Pass	11:02am

Temperature Tests

Test	Status	Time
FC1	Pass	11:02am
SRC	Pass	11:02am
DET	Pass	11:02am
BAR	Pass	11:02am
BT	Pass	11:02am

Blank Tests

Test	Status	Time
AIR	Pass	11:02am

Printer Tests

Test	Status	Time
PRNT	Pass	11:03am

CRC Tests

Test	Status	Time
COMP	Pass	11:03am
CAL	Pass	11:03am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Allen Ray Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: 008707
Test Date: 10/21/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011702
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:32pm
AIR BLK	.00	11:33pm
ACCY CHK	.08	11:34pm
AIR BLK	.00	11:34pm
SUB TEST	.00	11:35pm
AIR BLK	.00	11:36pm
SUB TEST	.00	11:37pm
AIR BLK	.00	11:38pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: 008707 Test Record Number: 690
Test Date: 10/21/2010 Test Time: 11:39pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:39pm
FLO	Pass	11:39pm
FC	Pass	11:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:39pm
SRC	Pass	11:39pm
DET	Pass	11:39pm
BAR	Pass	11:39pm
BT	Pass	11:39pm

Blank Tests

Test	Status	Time
AIR	Pass	11:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:40pm

CRC Tests

Test	Status	Time
COMP	Pass	11:40pm
CAL	Pass	11:40pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: 008647
Test Date: 10/21/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011703

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	11:33pm
AIR BLK	.00	11:34pm
ACCY CHK	.08	11:35pm
AIR BLK	.00	11:35pm
SUB TEST	.00	11:36pm
AIR BLK	.00	11:37pm
SUB TEST	.00	11:38pm
AIR BLK	.00	11:39pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: 008647 Test Record Number: 931
Test Date: 10/21/2010 Test Time: 11:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40pm
FLO	Pass	11:40pm
FC	Pass	11:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:40pm
SRC	Pass	11:40pm
DET	Pass	11:40pm
BAR	Pass	11:40pm
BT	Pass	11:40pm

Blank Tests

Test	Status	Time
AIR	Pass	11:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:41pm

CRC Tests

Test	Status	Time
COMP	Pass	11:41pm
CAL	Pass	11:41pm

Preventive Maintenance
Status: Pass

Alexis Kay Barnes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County MECKLENBURG Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008736 CHARLOTTE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY BAT MOBILE UNIT 3
590

Serial Number: 008736
Test Date: 10/21/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925102
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	11:31pm
AIR BLK	.00	11:32pm
ACCY CHK	.07	11:33pm
AIR BLK	.00	11:34pm
SUB TEST	.00	11:34pm
AIR BLK	.00	11:35pm
SUB TEST	.00	11:37pm
AIR BLK	.00	11:38pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY BAT MOBILE UNIT 3 590

Serial Number: 008736 Test Record Number: 210
Test Date: 10/21/2010 Test Time: 11:38pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:39pm
FLO	Pass	11:39pm
FC	Pass	11:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:39pm
SRC	Pass	11:39pm
DET	Pass	11:39pm
BAR	Pass	11:39pm
BT	Pass	11:39pm

Blank Tests

Test	Status	Time
AIR	Pass	11:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:39pm

CRC Tests

Test	Status	Time
COMP	Pass	11:40pm
CAL	Pass	11:40pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County IREDELL Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008707 STATESVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008707
Test Date: 10/23/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	8:53pm
AIR BLK	.00	8:54pm
ACCY CHK	.08	8:55pm
AIR BLK	.00	8:55pm
SUB TEST	.00	8:56pm
AIR BLK	.00	8:57pm
SUB TEST	.00	8:58pm
AIR BLK	.00	8:59pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008707 Test Record Number: 696
Test Date: 10/23/2010 Test Time: 8:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:00pm
FLO	Pass	9:00pm
FC	Pass	9:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:00pm
SRC	Pass	9:00pm
DET	Pass	9:00pm
BAR	Pass	9:00pm
BT	Pass	9:00pm

Blank Tests

Test	Status	Time
AIR	Pass	9:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:01pm

CRC Tests

Test	Status	Time
COMP	Pass	9:01pm
CAL	Pass	9:01pm

Preventive Maintenance
Status: Pass

Alvin Ray Banes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County IREDELL Instrument Location BAT MOBILE UNIT 3

Instrument Serial No. 008647 STATESVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Baines
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008647
Test Date: 10/23/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011703

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	8:55pm
AIR BLK	.00	8:56pm
ACCY CHK	.08	8:57pm
AIR BLK	.00	8:58pm
SUB TEST	.00	8:58pm
AIR BLK	.00	8:59pm
SUB TEST	.00	9:00pm
AIR BLK	.00	9:01pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008647 Test Record Number: 937
Test Date: 10/23/2010 Test Time: 9:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:02pm
FLO	Pass	9:02pm
FC	Pass	9:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:03pm
SRC	Pass	9:03pm
DET	Pass	9:03pm
BAR	Pass	9:03pm
BT	Pass	9:03pm

Blank Tests

Test	Status	Time
AIR	Pass	9:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:03pm

CRC Tests

Test	Status	Time
COMP	Pass	9:04pm
CAL	Pass	9:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County IREDELL Instrument Location BART MOBILE UNIT 3

Instrument Serial No. 008736 STATESVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008736
Test Date: 10/23/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG925102

Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	8:53pm
AIR BLK	.00	8:54pm
ACCY CHK	.07	8:55pm
AIR BLK	.00	8:56pm
SUB TEST	.00	8:56pm
AIR BLK	.00	8:57pm
SUB TEST	.00	8:59pm
AIR BLK	.00	8:59pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY BAT MOBILE UNIT 3 480

Serial Number: 008736 Test Record Number: 218
Test Date: 10/23/2010 Test Time: 9:00pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:00pm
FLO	Pass	9:00pm
FC	Pass	9:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:00pm
SRC	Pass	9:00pm
DET	Pass	9:00pm
BAR	Pass	9:00pm
BT	Pass	9:00pm

Blank Tests

Test	Status	Time
AIR	Pass	9:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:01pm

CRC Tests

Test	Status	Time
COMP	Pass	9:01pm
CAL	Pass	9:01pm

Preventive Maintenance
Status: Pass

Adam Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Durham Instrument Location BAT Mobile Unit 4

Instrument Serial No. 088734

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 4 310

Serial Number: 008734 Test Record Number: 391
Test Date: 10/23/2010 Test Time: 9:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:11pm
FLO	Pass	9:11pm
FC	Pass	9:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:11pm
SRC	Pass	9:11pm
DET	Pass	9:11pm
BAR	Pass	9:11pm
BT	Pass	9:11pm

Blank Tests

Test	Status	Time
AIR	Pass	9:12pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:12pm

CRC Tests

Test	Status	Time
COMP	Pass	9:12pm
CAL	Pass	9:12pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 4 310

Serial Number: 008734
Test Date: 10/23/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, - MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TRUDELL, SR., DANIEL T

Permit Number: 21535E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:03pm
AIR BLK	.00	9:04pm
ACCY CHK	.08	9:05pm
AIR BLK	.00	9:06pm
SUB TEST	.00	9:06pm
AIR BLK	.00	9:07pm
SUB TEST	.00	9:08pm
AIR BLK	.00	9:09pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Durham Instrument Location BAT Mobile Unit 4

Instrument Serial No. 008717

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 4 310

Serial Number: 008717 Test Record Number: 213
Test Date: 10/23/2010 Test Time: 10:56pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:57pm
FLO	Pass	10:57pm
FC	Pass	10:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:57pm
SRC	Pass	10:57pm
DET	Pass	10:57pm
BAR	Pass	10:57pm
BT	Pass	10:57pm

Blank Tests

Test	Status	Time
AIR	Pass	10:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:58pm

CRC Tests

Test	Status	Time
COMP	Pass	10:58pm
CAL	Pass	10:58pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 4 310

Serial Number: 008717
Test Date: 10/23/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TRUDELL, SR., DANIEL T

Permit Number: 21535E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG920302

Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	10:49pm
AIR BLK	.00	10:50pm
ACCY CHK	.08	10:50pm
AIR BLK	.00	10:51pm
SUB TEST	.00	10:52pm
AIR BLK	.00	10:53pm
SUB TEST	.00	10:54pm
AIR BLK	.00	10:55pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Lenoir Instrument Location Kinston P.D.

Instrument Serial No. 008918 205 E. King St., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY KINSTON PD 530

Serial Number: 008918
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG916602

Exp Date: 06/15/2011

Test	g/210L	Time
DIAG	Pass	2:46pm
AIR BLK	.00	2:47pm
ACCY CHK	.08	2:48pm
AIR BLK	.00	2:49pm
SUB TEST	.00	2:49pm
AIR BLK	.00	2:50pm
SUB TEST	.00	2:52pm
AIR BLK	.00	2:53pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY KINSTON PD 530

Serial Number: 008918 Test Record Number: 197
Test Date: 10/06/2010 Test Time: 2:54pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:54pm
FLO	Pass	2:54pm
FC	Pass	2:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:55pm
SRC	Pass	2:55pm
DET	Pass	2:55pm
BAR	Pass	2:55pm
BT	Pass	2:55pm

Blank Tests

Test	Status	Time
AIR	Pass	2:55pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:55pm

CRC Tests

Test	Status	Time
COMP	Pass	2:55pm
CAL	Pass	2:55pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Camden Instrument Location Camden Co. S.O.

Instrument Serial No. 008940 113 Hwy 343, Camden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of October, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly A. L. P.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG916602

Exp Date: 06/15/2011

Test	g/210L	Time
DIAG	Pass	11:58am
AIR BLK	.00	11:59am
ACCY CHK	.08	12:00pm
AIR BLK	.00	12:00pm
SUB TEST	.00	12:01pm
AIR BLK	.00	12:02pm
SUB TEST	.00	12:03pm
AIR BLK	.00	12:04pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940 Test Record Number: 297
Test Date: 10/04/2010 Test Time: 12:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:14pm
FLO	Pass	12:14pm
FC	Pass	12:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:14pm
SRC	Pass	12:14pm
DET	Pass	12:14pm
BAR	Pass	12:14pm
BT	Pass	12:14pm

Blank Tests

Test	Status	Time
AIR	Pass	12:15pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:15pm

CRC Tests

Test	Status	Time
COMP	Pass	12:15pm
CAL	Pass	12:15pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Richmond Instrument Location Richmond County
Instrument Serial No. 008840 MAGISTRATE'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 04 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul T. Guinn
Signature of Certifying Official

578
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RICHMOND COUNTY RICHMOND CO. MAG OFF
760

Serial Number: 008840
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SIMMONS, PAUL T

Permit Number: 08619E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011701

Exp Date: 04/27/2011

Test	g/210L	Time
DIAG	Pass	2:00pm
AIR BLK	.00	2:00pm
ACCY CHK	.08	2:01pm
AIR BLK	.00	2:02pm
SUB TEST	.00	2:03pm
AIR BLK	.00	2:04pm
SUB TEST	.00	2:05pm
AIR BLK	.00	2:06pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008840 Test Record Number: 516
Test Date: 10/04/2010 Test Time: 2:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:10pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ANSON Instrument Location ANSON County

Instrument Serial No. 008739 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 04 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul T. [Signature]
Signature of Certifying Official

578
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON CO. SO. 030

Serial Number: 008739
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SIMMONS, PAUL T
Permit Number: 08619E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG002802
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	12:39pm
AIR BLK	.00	12:39pm
ACCY CHK	.08	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	.00	12:41pm
AIR BLK	.00	12:42pm
SUB TEST	.00	12:44pm
AIR BLK	.00	12:45pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON CO. SO. 030

Serial Number: 008739 Test Record Number: 112
Test Date: 10/04/2010 Test Time: 12:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:46pm
FLO	Pass	12:46pm
FC	Pass	12:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:47pm
SRC	Pass	12:47pm
DET	Pass	12:47pm
BAR	Pass	12:47pm
BT	Pass	12:47pm

Blank Tests

Test	Status	Time
AIR	Pass	12:47pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:47pm

CRC Tests

Test	Status	Time
COMP	Pass	12:48pm
CAL	Pass	12:48pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ANSON Instrument Location ANSON County
Instrument Serial No. 008597 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 04 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul T. Quinn
Signature of Certifying Official

578
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

-Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SIMMONS, PAUL T
Permit Number: 08619E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG002802
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	12:36pm
AIR BLK	.00	12:37pm
ACCY CHK	.08	12:38pm
AIR BLK	.00	12:39pm
SUB TEST	.00	12:39pm
AIR BLK	.00	12:40pm
SUB TEST	.00	12:42pm
AIR BLK	.00	12:43pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597 Test Record Number: 466
Test Date: 10/04/2010 Test Time: 12:44pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:44pm
FLO	Pass	12:44pm
FC	Pass	12:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:45pm
SRC	Pass	12:45pm
DET	Pass	12:45pm
BAR	Pass	12:45pm
BT	Pass	12:45pm

Blank Tests

Test	Status	Time
AIR	Pass	12:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:45pm

CRC Tests

Test	Status	Time
COMP	Pass	12:45pm
CAL	Pass	12:45pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Richmond Instrument Location Richmond County

Instrument Serial No. 008701 MAGISTRATE'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 04 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul T. Johnson
Signature of Certifying Official

578
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RICHMOND COUNTY RICHMOND CO. MAG OFF
760

Serial Number: 008701
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SIMMONS, PAUL T
Permit Number: 08619E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG925102
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	2:01pm
AIR BLK	.00	2:02pm
ACCY CHK	.08	2:02pm
AIR BLK	.00	2:03pm
SUB TEST	.00	2:04pm
AIR BLK	.00	2:05pm
SUB TEST	.00	2:06pm
AIR BLK	.00	2:07pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008701 Test Record Number: 822
Test Date: 10/04/2010 Test Time: 2:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:10pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RANDOLPH Instrument Location BATMOBILE UNIT 3

Instrument Serial No. 008647 HIGH POINT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008647
Test Date: 10/09/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011703


Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	9:51pm
AIR BLK	.00	9:52pm
ACCY CHK	.08	9:52pm
AIR BLK	.00	9:53pm
SUB TEST	.00	9:54pm
AIR BLK	.00	9:55pm
SUB TEST	.00	9:56pm
AIR BLK	.00	9:57pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008647 Test Record Number: 922
Test Date: 10/09/2010 Test Time: 10:00pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:00pm
FLO	Pass	10:00pm
FC	Pass	10:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:01pm
SRC	Pass	10:01pm
DET	Pass	10:01pm
BAR	Pass	10:01pm
BT	Pass	10:01pm

Blank Tests

Test	Status	Time
AIR	Pass	10:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:01pm

CRC Tests

Test	Status	Time
COMP	Pass	10:01pm
CAL	Pass	10:01pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RAWDOLPH Instrument Location BAT MOBILE UNIT 3
Instrument Serial No. 008707 HIGH POINT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alexis Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008707

Test Date: 10/09/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	9:48pm
AIR BLK	.00	9:49pm
ACCY CHK	.08	9:50pm
AIR BLK	.00	9:51pm
SUB TEST	.00	9:51pm
AIR BLK	.00	9:52pm
SUB TEST	.00	9:54pm
AIR BLK	.00	9:55pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008707 Test Record Number: 682
Test Date: 10/09/2010 Test Time: 9:55pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55pm
FLO	Pass	9:55pm
FC	Pass	9:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:56pm
SRC	Pass	9:56pm
DET	Pass	9:56pm
BAR	Pass	9:56pm
BT	Pass	9:56pm

Blank Tests

Test	Status	Time
AIR	Pass	9:56pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:56pm

CRC Tests

Test	Status	Time
COMP	Pass	9:57pm
CAL	Pass	9:57pm

Preventive Maintenance
Status: Pass

Alun Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RANDOLPH Instrument Location BAT MOBILE UNIT 3
Instrument Serial No. 008736 HIGH POINT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Allen Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008736
Test Date: 10/09/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925102
Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	9:48pm
AIR BLK	.00	9:49pm
ACCY CHK	.07	9:50pm
AIR BLK	.00	9:50pm
SUB TEST	.00	9:51pm
AIR BLK	.00	9:52pm
SUB TEST	.00	9:53pm
AIR BLK	.00	9:54pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY BAT MOBILE UNIT 3 750

Serial Number: 008736 Test Record Number: 200
Test Date: 10/09/2010 Test Time: 9:55pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:55pm
FLO	Pass	9:55pm
FC	Pass	9:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:55pm
SRC	Pass	9:55pm
DET	Pass	9:55pm
BAR	Pass	9:55pm
BT	Pass	9:55pm

Blank Tests

Test	Status	Time
AIR	Pass	9:56pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:56pm

CRC Tests

Test	Status	Time
COMP	Pass	9:56pm
CAL	Pass	9:56pm

Preventive Maintenance
Status: Pass

Allen Ray Barnes

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location BURLINGTON P.D.

Instrument Serial No. 008812 267 W. FRONT ST. BURLINGTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008812
Test Date: 10/08/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG002803

Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:28am
AIR BLK	.00	11:29am
ACCY CHK	.07	11:30am
AIR BLK	.00	11:31am
SUB TEST	.00	11:31am
AIR BLK	.00	11:32am
SUB TEST	.00	11:34am
AIR BLK	.00	11:35am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008812 Test Record Number: 693
Test Date: 10/08/2010 Test Time: 11:38am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:38am

Temperature Tests

Test	Status	Time
FC1	Pass	11:38am
SRC	Pass	11:38am
DET	Pass	11:38am
BAR	Pass	11:38am
BT	Pass	11:38am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am


Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:39am
CAL	Pass	11:39am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location BURLINGTON P.D.

Instrument Serial No. 008907 267 W. FRONT ST. BURLINGTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of OCTOBER, 200 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008907
Test Date: 10/08/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002803
Exp Date: 01/28/2012

Test	g/210L	Time
DIAG	Pass	11:32am
AIR BLK	.00	11:33am
ACCY CHK	.08	11:34am
AIR BLK	.00	11:35am
SUB TEST	.00	11:35am
AIR BLK	.00	11:36am
SUB TEST	.00	11:38am
AIR BLK	.00	11:39am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008907 Test Record Number: 340
Test Date: 10/08/2010 Test Time: 11:41am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:41am
FLO	Pass	11:41am
FC	Pass	11:41am

Temperature Tests

Test	Status	Time
FC1	Pass	11:42am
SRC	Pass	11:42am
DET	Pass	11:42am
BAR	Pass	11:42am
BT	Pass	11:42am

Blank Tests

Test	Status	Time
AIR	Pass	11:42am

Printer Tests

Test	Status	Time
PRNT	Pass	11:42am

CRC Tests

Test	Status	Time
COMP	Pass	11:42am
CAL	Pass	11:42am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO. JAIL

Instrument Serial No. 008913 109 S. MAPLE ST. GRAHAM, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913
Test Date: 10/08/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011703

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	12:55pm
AIR BLK	.00	12:56pm
ACCY CHK	.07	12:57pm
AIR BLK	.00	12:58pm
SUB TEST	.00	12:58pm
AIR BLK	.00	12:59pm
SUB TEST	.00	1:01pm
AIR BLK	.00	1:02pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913 Test Record Number: 894
Test Date: 10/08/2010 Test Time: 1:03pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:03pm
FLO	Pass	1:03pm
FC	Pass	1:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:03pm
SRC	Pass	1:03pm
DET	Pass	1:03pm
BAR	Pass	1:03pm
BT	Pass	1:03pm

Blank Tests

Test	Status	Time
AIR	Pass	1:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:04pm

CRC Tests

Test	Status	Time
COMP	Pass	1:04pm
CAL	Pass	1:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO. TAN

Instrument Serial No. 008924 109 S. MAPLE ST. GRATHAM, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008924
Test Date: 10/08/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG016501
Exp Date: 06/14/2012


Test	g/210L	Time
DIAG	Pass	2:23pm
AIR BLK	.00	2:24pm
ACCY CHK	.08	2:25pm
AIR BLK	.00	2:26pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:29pm
AIR BLK	.00	2:29pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008924 Test Record Number: 69
Test Date: 10/08/2010 Test Time: 2:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:31pm
FLO	Pass	2:31pm
FC	Pass	2:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:31pm
SRC	Pass	2:31pm
DET	Pass	2:31pm
BAR	Pass	2:31pm
BT	Pass	2:31pm

Blank Tests

Test	Status	Time
AIR	Pass	2:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:32pm

CRC Tests

Test	Status	Time
COMP	Pass	2:32pm
CAL	Pass	2:32pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GRANVILLE Instrument Location OXFORD P.D.

Instrument Serial No. 008923 204 E. McCLANAHAN ST. OXFORD, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



N. J. [Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY OXFORD PD 380

Serial Number: 008923
Test Date: 10/11/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902603

Exp Date: 01/26/2011


Test	g/210L	Time
DIAG	Pass	12:25pm
AIR BLK	.00	12:26pm
ACCY CHK	.07	12:27pm
AIR BLK	.00	12:28pm
SUB TEST	.00	12:28pm
AIR BLK	.00	12:29pm
SUB TEST	.00	12:31pm
AIR BLK	.00	12:32pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY OXFORD PD 380

Serial Number: 008923 Test Record Number: 338
Test Date: 10/11/2010 Test Time: 12:33pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:33pm
FLO	Pass	12:33pm
FC	Pass	12:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:33pm
SRC	Pass	12:33pm
DET	Pass	12:33pm
BAR	Pass	12:33pm
BT	Pass	12:33pm

Blank Tests

Test	Status	Time
AIR	Pass	12:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:34pm

CRC Tests

Test	Status	Time
COMP	Pass	12:34pm
CAL	Pass	12:34pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County SRAVILLE Instrument Location CREEDMOOR P.D.

Instrument Serial No. 008641 111 MASONIC ST. CREEDMOOR, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641
Test Date: 10/11/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG925102
Exp Date: 09/08/2011


Test	g/210L	Time
DIAG	Pass	1:13pm
AIR BLK	.00	1:14pm
ACCY CHK	.07	1:15pm
AIR BLK	.00	1:15pm
SUB TEST	.00	1:16pm
AIR BLK	.00	1:17pm
SUB TEST	.00	1:18pm
AIR BLK	.00	1:19pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641 Test Record Number: 560
Test Date: 10/11/2010 Test Time: 1:20pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:20pm
FLO	Pass	1:20pm
FC	Pass	1:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:20pm
SRC	Pass	1:20pm
DET	Pass	1:20pm
BAR	Pass	1:20pm
BT	Pass	1:20pm

Blank Tests

Test	Status	Time
AIR	Pass	1:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:21pm

CRC Tests

Test	Status	Time
COMP	Pass	1:21pm
CAL	Pass	1:21pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location B&T Mobile Unit 95
Instrument Serial No. 008600 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Steph G. Thompson
Signature of Certifying Official 636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008600 Test Record Number: 682
Test Date: 10/09/2010 Test Time: 9:45pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:46pm
FLO	Pass	9:46pm
FC	Pass	9:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:46pm
SRC	Pass	9:46pm
DET	Pass	9:46pm
BAR	Pass	9:46pm
BT	Pass	9:46pm

Blank Tests

Test	Status	Time
AIR	Pass	9:47pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:47pm

CRC Tests

Test	Status	Time
COMP	Pass	9:47pm
CAL	Pass	9:47pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008600
Test Date: 10/09/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G
Permit Number: 09372E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG920302
Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:37pm
AIR BLK	.00	9:38pm
ACCY CHK	.08	9:39pm
AIR BLK	.00	9:40pm
SUB TEST	.00	9:40pm
AIR BLK	.00	9:41pm
SUB TEST	.00	9:43pm
AIR BLK	.00	9:44pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake

Instrument Location BAT MOBILE UNIT #5


Instrument Serial No. 008698

HOLLY SPRINGS

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Steph G. Thomas 1636
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698 Test Record Number: 564
Test Date: 10/09/2010 Test Time: 9:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:47pm
FLO	Pass	9:47pm
FC	Pass	9:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:47pm
SRC	Pass	9:47pm
DET	Pass	9:47pm
BAR	Pass	9:47pm
BT	Pass	9:47pm

Blank Tests

Test	Status	Time
AIR	Pass	9:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:48pm

CRC Tests

Test	Status	Time
COMP	Pass	9:48pm
CAL	Pass	9:48pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008698
Test Date: 10/09/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG920302


Exp Date: 07/22/2011

Test	g/210L	Time
DIAG	Pass	9:38pm
AIR BLK	.00	9:39pm
ACCY CHK	.08	9:40pm
AIR BLK	.00	9:41pm
SUB TEST	.00	9:41pm
AIR BLK	.00	9:42pm
SUB TEST	.00	9:43pm
AIR BLK	.00	9:44pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location BAT MOBILE UNIT #5
Instrument Serial No. 008788 Holly Springs

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Steph G. Thompson 636
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788 Test Record Number: 467
Test Date: 10/09/2010 Test Time: 9:47pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:48pm
FLO	Pass	9:48pm
FC	Pass	9:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:48pm
SRC	Pass	9:48pm
DET	Pass	9:48pm
BAR	Pass	9:48pm
BT	Pass	9:48pm

Blank Tests

Test	Status	Time
AIR	Pass	9:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:49pm

CRC Tests

Test	Status	Time
COMP	Pass	9:49pm
CAL	Pass	9:49pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 5 910

Serial Number: 008788

Test Date: 10/09/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 09372E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904903


Exp Date: 02/18/2011

Test	g/210L	Time
DIAG	Pass	9:39pm
AIR BLK	.00	9:40pm
ACCY CHK	.08	9:41pm
AIR BLK	.00	9:41pm
SUB TEST	.00	9:42pm
AIR BLK	.00	9:43pm
SUB TEST	.00	9:44pm
AIR BLK	.00	9:45pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DURHAM Instrument Location DURHAM CO. JAIL

Instrument Serial No. 008891 217 S. MANGUM ST. DURHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Brian D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008891
Test Date: 10/12/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 08937E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG002602

Exp Date: 01/26/2012

Test	g/210L	Time
DIAG	Pass	1:42pm
AIR BLK	.00	1:43pm
ACCY CHK	.08	1:44pm
AIR BLK	.00	1:44pm
SUB TEST	.00	1:45pm
AIR BLK	.00	1:46pm
SUB TEST	.00	1:48pm
AIR BLK	.00	1:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008891 Test Record Number: 899
Test Date: 10/12/2010 Test Time: 1:51pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:51pm
FLO	Pass	1:51pm
FC	Pass	1:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:51pm
SRC	Pass	1:51pm
DET	Pass	1:51pm
BAR	Pass	1:51pm
BT	Pass	1:51pm

Blank Tests

Test	Status	Time
AIR	Pass	1:52pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:52pm

CRC Tests

Test	Status	Time
COMP	Pass	1:52pm
CAL	Pass	1:52pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Durham Instrument Location Durham Co. Jail

Instrument Serial No. 008878 217 S. Mangum St. Durham, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008878

Test Date: 10/12/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011701


Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	1:42pm
AIR BLK	.00	1:43pm
ACCY CHK	.08	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	.00	1:45pm
AIR BLK	.00	1:46pm
SUB TEST	.00	1:48pm
AIR BLK	.00	1:48pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008878 Test Record Number: 1118
Test Date: 10/12/2010 Test Time: 1:52pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:52pm
FLO	Pass	1:52pm
FC	Pass	1:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:53pm
SRC	Pass	1:53pm
DET	Pass	1:53pm
BAR	Pass	1:53pm
BT	Pass	1:53pm

Blank Tests

Test	Status	Time
AIR	Pass	1:53pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:53pm

CRC Tests

Test	Status	Time
COMP	Pass	1:53pm
CAL	Pass	1:53pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DURHAM Instrument Location DURHAM CO. JAIL

Instrument Serial No. 008859 217 S. MANGUM ST. DURHAM, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008859
Test Date: 10/12/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG002602

Exp Date: 01/26/2012

Test	g/210L	Time
DIAG	Pass	1:41pm
AIR BLK	.00	1:42pm
ACCY CHK	.08	1:43pm
AIR BLK	.00	1:44pm
SUB TEST	.00	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	.00	1:47pm
AIR BLK	.00	1:48pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008859 Test Record Number: 657
Test Date: 10/12/2010 Test Time: 1:49pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:49pm
FLO	Pass	1:49pm
FC	Pass	1:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:49pm
SRC	Pass	1:49pm
DET	Pass	1:49pm
BAR	Pass	1:49pm
BT	Pass	1:49pm

Blank Tests

Test	Status	Time
AIR	Pass	1:50pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:50pm

CRC Tests

Test	Status	Time
COMP	Pass	1:50pm
CAL	Pass	1:50pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County JOHNSTON Instrument Location SELMA POLICE DEPT.

Instrument Serial No. 008595 SELMA NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 06 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 06108E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG003403
Exp Date: 02/03/2012

Test	g/210L	Time
DIAG	Pass	2:51pm
AIR BLK	.00	2:51pm
ACCY CHK	.07	2:52pm
AIR BLK	.00	2:53pm
SUB TEST	.00	2:54pm
AIR BLK	.00	2:55pm
SUB TEST	.00	2:56pm
AIR BLK	.00	2:57pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595 Test Record Number: 443
Test Date: 10/06/2010 Test Time: 2:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:00pm
FLO	Pass	3:00pm
FC	Pass	3:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:00pm
SRC	Pass	3:00pm
DET	Pass	3:00pm
BAR	Pass	3:00pm
BT	Pass	3:00pm

Blank Tests

Test	Status	Time
AIR	Pass	3:01pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:01pm

CRC Tests

Test	Status	Time
COMP	Pass	3:01pm
CAL	Pass	3:01pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Haywood Instrument Location Port Mobile Unit 4

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dennis O. Thomas
Signature of Certifying Official

651
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY BAT MOBILE UNIT 4 430

Serial Number: 008871 Test Record Number: 317
Test Date: 10/09/2010 Test Time: 5:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:37pm
FLO	Pass	5:37pm
FC	Pass	5:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:37pm
SRC	Pass	5:37pm
DET	Pass	5:37pm
BAR	Pass	5:37pm
BT	Pass	5:37pm

Blank Tests

Test	Status	Time
AIR	Pass	5:38pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:38pm

CRC Tests

Test	Status	Time
COMP	Pass	5:38pm
CAL	Pass	5:38pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

HAYWOOD COUNTY BAT MOBILE UNIT 4 430

Serial Number: 008871
Test Date: 10/09/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TRUDELL, SR., DANIEL T

Permit Number: 21535E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011703


Exp Date: 10/02/2011

Test	g/210L	Time
DIAG	Pass	5:30pm
AIR BLK	.00	5:31pm
ACCY CHK	.08	5:31pm
AIR BLK	.00	5:32pm
SUB TEST	.00	5:32pm
AIR BLK	.00	5:33pm
SUB TEST	.00	5:35pm
AIR BLK	.00	5:36pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CARY P.D.

Instrument Serial No. 008587 120 WILKINSON AVE. CARY, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CARY PD 910

Serial Number: 008587
Test Date: 10/07/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG011703


Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	10:36am
AIR BLK	.00	10:37am
ACCY CHK	.08	10:38am
AIR BLK	.00	10:39am
SUB TEST	.00	10:39am
AIR BLK	.00	10:40am
SUB TEST	.00	10:41am
AIR BLK	.00	10:42am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CARY PD 910

Serial Number: 008587 Test Record Number: 1195
Test Date: 10/07/2010 Test Time: 10:44am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:45am
FLO	Pass	10:45am
FC	Pass	10:45am

Temperature Tests

Test	Status	Time
FC1	Pass	10:45am
SRC	Pass	10:45am
DET	Pass	10:45am
BAR	Pass	10:45am
BT	Pass	10:45am

Blank Tests

Test	Status	Time
AIR	Pass	10:45am


Printer Tests

Test	Status	Time
PRNT	Pass	10:46am

CRC Tests

Test	Status	Time
COMP	Pass	10:46am
CAL	Pass	10:46am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ORANGE Instrument Location CHAPEL HILL P.D.

Instrument Serial No. 008839 828 MARTIN LUTHER KING JR. BLD
CHAPEL HILL, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839
Test Date: 10/06/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG016501

Exp Date: 06/14/2012


Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.08	11:32am
AIR BLK	.00	11:33am
SUB TEST	.00	11:34am
AIR BLK	.00	11:35am
SUB TEST	.00	11:36am
AIR BLK	.00	11:37am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839 Test Record Number: 582
Test Date: 10/06/2010 Test Time: 11:39am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40am
FLO	Pass	11:40am
FC	Pass	11:40am

Temperature Tests

Test	Status	Time
FC1	Pass	11:40am
SRC	Pass	11:40am
DET	Pass	11:40am
BAR	Pass	11:40am
BT	Pass	11:40am

Blank Tests

Test	Status	Time
AIR	Pass	11:40am


Printer Tests

Test	Status	Time
PRNT	Pass	11:40am

CRC Tests

Test	Status	Time
COMP	Pass	11:41am
CAL	Pass	11:41am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ORANGE Instrument Location CHAPEL HILL P.D.

Instrument Serial No. 008856 826 MARTIN LUTHER KING JR. BLVD
CHAPEL HILL, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008856
Test Date: 10/06/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002602
Exp Date: 01/26/2012


Test	g/210L	Time
DIAG	Pass	11:38am
AIR BLK	.00	11:39am
ACCY CHK	.08	11:39am
AIR BLK	.00	11:40am
SUB TEST	.00	11:41am
AIR BLK	.00	11:42am
SUB TEST	.00	11:43am
AIR BLK	.00	11:44am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008856 Test Record Number: 576
Test Date: 10/06/2010 Test Time: 11:45am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:46am
FLO	Pass	11:46am
FC	Pass	11:46am

Temperature Tests

Test	Status	Time
FC1	Pass	11:46am
SRC	Pass	11:46am
DET	Pass	11:46am
BAR	Pass	11:46am
BT	Pass	11:46am

Blank Tests

Test	Status	Time
AIR	Pass	11:47am


Printer Tests

Test	Status	Time
PRNT	Pass	11:47am

CRC Tests

Test	Status	Time
COMP	Pass	11:47am
CAL	Pass	11:47am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County CASWELL Instrument Location SMP CASWELL OFFICE

Instrument Serial No. 008593 956 FIRETOWER RD, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CASWELL COUNTY SHP YANCEYVILLE 160

Serial Number: 008593
Test Date: 10/05/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG002602

Exp Date: 01/26/2012

Test	g/210L	Time
DIAG	Pass	11:06am
AIR BLK	.00	11:07am
ACCY CHK	.08	11:07am
AIR BLK	.00	11:08am
SUB TEST	.00	11:09am
AIR BLK	.00	11:09am
SUB TEST	.00	11:11am
AIR BLK	.00	11:12am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CASWELL COUNTY SHP YANCEYVILLE 160

Serial Number: 008593 Test Record Number: 597
Test Date: 10/05/2010 Test Time: 11:13am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:13am
FLO	Pass	11:13am
FC	Pass	11:13am

Temperature Tests

Test	Status	Time
FC1	Pass	11:13am
SRC	Pass	11:13am
DET	Pass	11:13am
BAR	Pass	11:13am
BT	Pass	11:13am

Blank Tests

Test	Status	Time
AIR	Pass	11:14am


Printer Tests

Test	Status	Time
PRNT	Pass	11:14am

CRC Tests

Test	Status	Time
COMP	Pass	11:14am
CAL	Pass	11:14am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CCRD

Instrument Serial No. 008816 330 S SALISBURY ST. RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CCBI 910

Serial Number: 008816
Test Date: 10/04/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLA, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011701

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	10:32am
AIR BLK	.00	10:33am
ACCY CHK	.08	10:33am
AIR BLK	.00	10:34am
SUB TEST	.00	10:35am
AIR BLK	.00	10:36am
SUB TEST	.00	10:37am
AIR BLK	.00	10:38am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CCBI 910

Serial Number: 008816 Test Record Number: 3388
Test Date: 10/04/2010 Test Time: 10:40am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:41am
FLO	Pass	10:41am
FC	Pass	10:41am

Temperature Tests

Test	Status	Time
FC1	Pass	10:41am
SRC	Pass	10:41am
DET	Pass	10:41am
BAR	Pass	10:41am
BT	Pass	10:41am

Blank Tests

Test	Status	Time
AIR	Pass	10:42am


Printer Tests

Test	Status	Time
PRNT	Pass	10:42am

CRC Tests

Test	Status	Time
COMP	Pass	10:42am
CAL	Pass	10:42am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CCBI

Instrument Serial No. 008686 330 S. SALISBURY ST. RALEIGH, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 20 10 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CCBI 910

Serial Number: 008686
Test Date: 10/04/2010

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name:

QUARANTELLLO, NICHOLAS J

Permit Number: 21536E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG002602

Exp Date: 01/26/2012


Test	g/210L	Time
DIAG	Pass	10:06am
AIR BLK	.00	10:08am
ACCY CHK	.07	10:08am
AIR BLK	.00	10:09am
SUB TEST	.00	10:10am
AIR BLK	.00	10:11am
SUB TEST	.00	10:12am
AIR BLK	.00	10:13am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CCBI 910

Serial Number: 008686 Test Record Number: 2262
Test Date: 10/04/2010 Test Time: 10:15am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:16am
FLO	Pass	10:16am
FC	Pass	10:16am

Temperature Tests

Test	Status	Time
FC1	Pass	10:16am
SRC	Pass	10:16am
DET	Pass	10:16am
BAR	Pass	10:16am
BT	Pass	10:16am

Blank Tests

Test	Status	Time
AIR	Pass	10:17am


Printer Tests

Test	Status	Time
PRNT	Pass	10:17am

CRC Tests

Test	Status	Time
COMP	Pass	10:17am
CAL	Pass	10:17am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location KNIGHTDALE P.S.

Instrument Serial No. 008838 979 STEEPIS SQUARE CT. KNIGHTDALE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

652
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY KNIGHTDALE PS 910

Serial Number: 008838
Test Date: 10/04/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name:
QUARANTELLA, NICHOLAS J
Permit Number: 21536E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG002602
Exp Date: 01/26/2012

Test	g/210L	Time
DIAG	Pass	11:45am
AIR BLK	.00	11:46am
ACCY CHK	.08	11:46am
AIR BLK	.00	11:47am
SUB TEST	.00	11:48am
AIR BLK	.00	11:49am
SUB TEST	.00	11:50am
AIR BLK	.00	11:51am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY. KNIGHTDALE PS 910

Serial Number: 008838 Test Record Number: 372
Test Date: 10/04/2010 Test Time: 11:52am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:53am
FLO	Pass	11:53am
FC	Pass	11:53am

Temperature Tests

Test	Status	Time
FC1	Pass	11:53am
SRC	Pass	11:53am
DET	Pass	11:53am
BAR	Pass	11:53am
BT	Pass	11:53am

Blank Tests

Test	Status	Time
AIR	Pass	11:53am

Printer Tests

Test	Status	Time
PRNT	Pass	11:53am

CRC Tests

Test	Status	Time
COMP	Pass	11:54am
CAL	Pass	11:54am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location BATMOBILE UNIT 3

Instrument Serial No. 008707 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008707
Test Date: 10/01/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG011702

Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	8:04pm
AIR BLK	.00	8:05pm
ACCY CHK	.08	8:05pm
AIR BLK	.00	8:06pm
SUB TEST	.00	8:06pm
AIR BLK	.00	8:07pm
SUB TEST	.00	8:09pm
AIR BLK	.00	8:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008707 Test Record Number: 677
Test Date: 10/01/2010 Test Time: 8:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:10pm
FLO	Pass	8:10pm
FC	Pass	8:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:11pm
SRC	Pass	8:11pm
DET	Pass	8:11pm
BAR	Pass	8:11pm
BT	Pass	8:11pm

Blank Tests

Test	Status	Time
AIR	Pass	8:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:11pm

CRC Tests

Test	Status	Time
COMP	Pass	8:11pm
CAL	Pass	8:11pm

Preventive Maintenance
Status: Pass

Alan Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location BATMOBILE UNIT 3

Instrument Serial No. 008647 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008647
Test Date: 10/01/2010

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
10/01/2009-10/01/2011

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG011703
Exp Date: 04/27/2012

Test	g/210L	Time
DIAG	Pass	8:04pm
AIR BLK	.00	8:05pm
ACCY CHK	.08	8:06pm
AIR BLK	.00	8:06pm
SUB TEST	.00	8:07pm
AIR BLK	.00	8:08pm
SUB TEST	.00	8:09pm
AIR BLK	.00	8:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008647 Test Record Number: 919
Test Date: 10/01/2010 Test Time: 8:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:11pm
FLO	Pass	8:11pm
FC	Pass	8:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:11pm
SRC	Pass	8:11pm
DET	Pass	8:11pm
BAR	Pass	8:11pm
BT	Pass	8:11pm

Blank Tests

Test	Status	Time
AIR	Pass	8:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:12pm

CRC Tests

Test	Status	Time
COMP	Pass	8:12pm
CAL	Pass	8:12pm

Preventive Maintenance
Status: Pass

Alan Ray Bantz
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location BATMOBILE UNIT 3

Instrument Serial No. 008736 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of OCTOBER, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008736
Test Date: 10/01/2010

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

10/01/2009-10/01/2011

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG925102

Exp Date: 09/08/2011

Test	g/210L	Time
DIAG	Pass	8:05pm
AIR BLK	.00	8:06pm
ACCY CHK	.07	8:06pm
AIR BLK	.00	8:07pm
SUB TEST	.00	8:08pm
AIR BLK	.00	8:09pm
SUB TEST	.00	8:10pm
AIR BLK	.00	8:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY BAT MOBILE UNIT 3 400

Serial Number: 008736 Test Record Number: 197
Test Date: 10/01/2010 Test Time: 8:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:12pm
FLO	Pass	8:12pm
FC	Pass	8:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:12pm
SRC	Pass	8:12pm
DET	Pass	8:12pm
BAR	Pass	8:12pm
BT	Pass	8:12pm

Blank Tests

Test	Status	Time
AIR	Pass	8:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:13pm

CRC Tests

Test	Status	Time
COMP	Pass	8:13pm
CAL	Pass	8:13pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst